



SERVOPRO NANO Trace DF-745SGMAX Analyzer Installation and Operator Manual

Part Number: **225487**
Revision: **073122**
Language: **US English**



SERVOMEX.COM

IMPORTANT INFORMATION



Continued safe and reliable operation of this equipment is conditional on all installation, operation and maintenance procedures being carried out in accordance with the appropriate manuals, by personnel having appropriate qualifications, experience, and training.

Failure to observe the requirements of the manual may result in the user being held responsible for the consequences and may invalidate any warranty.

Servomex will accept no liability for unauthorized modifications to Servomex supplied equipment.

Servomex has paid particular attention to Health and Safety throughout this manual. Where special precautions need to be taken due to the nature of the equipment or product, an appropriate safety icon and warning message is shown. Special attention should be made to the Safety Chapter if available, where all such messages are summarized.

The user is solely responsible for implementing appropriate environmental monitoring, ventilation, and gas safety controls for flammable and/or toxic gas installations to meet all relevant safety standards including but not limited to those imposed by legal, governmental, statutory, industrial, regulatory and/or corporate requirements.

This instrument is not intended for use in applications to which a flammable gas is blended with an enriched oxygen background gas.

In line with our continuous policy of research and development, we reserve the right to amend models and specifications without prior notice.

This handbook is accurate at the date of printing but will be superseded and should be disregarded if specifications or appearance are changed.

Servomex is a registered trademark of Servomex Group Limited. The use of all trademarks in this document is acknowledged.

Table of Contents

1. Safety Summary	1
1.1 General Warnings and Cautions	1
1.2 Laser Safety	3
1.3 Safety Labelling and General Marking	4
2. Introduction	6
2.1 Product Description	6
2.1.1 Product Dimensions	6
2.1.2 Internal Components	8
2.2 Specifications Summary	9
2.2.1 Moisture Detection	9
2.2.2 Laser	10
2.2.3 Environmental Operating Conditions	10
3. Installation, Start-up, and Shutdown	11
3.1 Unpacking Procedure	11
3.2 Analyzer Installation Overview	12
3.3 Analyzer Installation	12
3.4 Analyzer Start-up	17
3.4.1 Gas Flow and Pressure Regulation	17
3.4.2 Gas Flow Setting (without Isolation Valves)	17
3.4.3 Gas Flow Setting (with Isolation Valves Option)	18
3.4.4 Back Flow Prevention System (Isolation Valves Option)	18
3.4.5 Download System Data	20
3.5 Analyzer Shut Down	20
4. Options	23
4.1 Key Lock	23
4.2 Operating Voltage	23
4.3 Serial Communications	23
4.4 Analog Voltage Output	23
4.5 Isolation Valves	23
4.6 Hydrogen Safety System	23
5. Connecting to External Devices	24
5.1 Serial Communication Port – J5	24
5.2 Analog Signal Outputs – J4	25
5.3 4-20 mA Outputs – J4	25
5.4 Relay Ports – J8, J9	26
6. User Interface	27
6.1 Data Display Screen	27
6.2 Keypad	27
6.3 Menu Structure	28
6.3.1 Main Menu	29
6.3.2 Set Up GSF	30
6.3.3 HFC Samples	35
6.3.4 Isolate Analyzer (Isolation Valves Option)	37
6.3.5 Restore Sample Gas Flow (Isolation Valves Option)	40

6.3.6	Calibrate Menu	40
6.3.7	Data History Routine	44
6.3.8	Data Downloader Routine	47
6.3.9	View Logs Menu	50
6.3.10	Analyzer Setup	55
6.3.11	Analog Output Setup	60
6.3.12	Graph Setup	61
6.3.13	Scaler Editor	63
6.3.14	Scheduled Reboot	64
6.3.15	Diagnostics Menu	65
6.3.16	Power Up Default (applicable to analyzers with iso valves)	76
6.3.17	Date/Time	78
6.3.18	Communications	80
6.3.19	Download System Data	81
6.3.20	System Info	83
7.	Sample Gas Preparation and Delivery	86
7.1	Introduction	86
7.2	Sample Pressure	86
7.3	Flow Rate Effects on Sensor Performance	86
7.4	Background Gas Effects on Indicated Flow Rate	86
7.5	Flammable Sample Gas	86
8.	Service	87
8.1	Return Material Authorization number	87
8.2	Cautions Related to Maintenance	87
8.3	Maintenance	87
8.3.1	Storage	87
8.3.2	Fuse Replacement	88
8.3.3	Moisture Cell Maintenance	88
8.3.4	Vacuum Pump Maintenance	88
8.3.5	Power Cord Replacement	88
8.4	Troubleshooting Guide for the NanoTrace Analyzer	89
8.4.1	Moisture Troubleshooting	89
8.5	Shipping	89
8.6	Disposal	90
8.6.1	Disposal in accordance with Waste Electrical and Electronic Equipment (WEEE)	90
9.	Theory of Operation	91
9.1	The Moisture Measurement	91
9.1.1	Moisture and the IR Spectrum	91
9.2	Absorption Spectroscopy	92
10.	Specifications	93
10.1	Moisture	93
10.2	General	93
10.3	Environmental:	93
10.4	Gas Routing	93
10.5	Output Signals & Status Indicators	94
10.6	Laser	94

10.8	Power.....	95
10.9	Compliance	95
11.	Warranty.....	96
	Index	97
	Appendix A – Hydrogen Safety System	99
	Instrument	99
	Vacuum Pump.....	99
	Installation	99
	Operation.....	100
	Appendix B: NanoTrace DF-700 Series Moisture Analyzers Serial Digital Communication Protocol	
	Programming Guide	103

Table of Figures

Figure 1: Analyzer Markings Shown in Rear Area View	4
Figure 2: Sensor Markings – Side	4
Figure 3: Sensor Marking – Front	5
Figure 4: DF-745SGMAX Analyzer (front)	6
Figure 5: DF-745SGMAX Analyzer (rear).....	7
Figure 6: DF-745SGMAX Analyzer (left)	7
Figure 7: Major Internal Components	9
Figure 8: Backup Wrench Use	12
Figure 9: VCR Gasket (with supplied mounting bracket)	13
Figure 10: DF-745SGMAX Analyzer Gas Panel	14
Figure 11: Starting Gas Flow	16
Figure 12: Data Display Screen	17
Figure 13: Back Flow Prevention Warning.....	19
Figure 14: Re-established flow delay	20
Figure 15: Block Diagram of Gas Flow Path	21
Figure 16: Block Diagram of Gas Flow Path with Optional Valves.....	21
Figure 17: Rear Panel Electrical Connectors	24
Figure 18: Data Display Screen	27
Figure 19: Keypad	28
Figure 20: Main Menu (Without Isolation Valves Option).....	29
Figure 21: Main Menu (With Isolation Valves Option)	30
Figure 22: Setup GSF Menu	31
Figure 23: Sample GSF Setup Screen	32
Figure 24: GSF Editor with Template Active	33
Figure 25: HFC Samples Menu Selection	35
Figure 26: Interfering Sample Gas Selection	36
Figure 27: Baseline Status Action Dialog Box	36
Figure 28: Baseline Data Screen	37
Figure 29: Isolate Analyzer.....	37
Figure 30: Isolate Warning	38
Figure 31: Back Flow Prevention Warning.....	39
Figure 32: Re-established flow delay	39
Figure 33: Restore Sample Gas Flow	40
Figure 34: Calibrate Menu	41
Figure 35: Check/Adjust Zero Menu	42
Figure 36: Check/Adjust Zero Screen.....	42
Figure 37: Manual Zero Screen.....	44
Figure 38: Data History Menu.....	45
Figure 39: Data History Screen	46
Figure 40: Install Media	46
Figure 41: Example of Data Download	47
Figure 42: Data Downloader Menu	47
Figure 43: Data Downloader Screen	48
Figure 44: View Location Screen.....	49
Figure 45: Keyboard Display	50
Figure 46: Delete Selection.....	50

Figure 47: View Logs Menu Selection	51
Figure 48: View Zero Log Menu	52
Figure 49: Zero Log Screen	52
Figure 50: View System Error Code Log Menu	53
Figure 51: System Error Log Screen	53
Figure 52: Fan Failure Alarm	54
Figure 53: View GSF Log Menu Selection	55
Figure 54: GSF Log	55
Figure 55: Alarm Setup Menu Selection	56
Figure 56: Alarm Setup Screen	57
Figure 57: Temperature Alarm Setup	58
Figure 58: Pressure Range Alarm Setup	59
Figure 59: System Alarm Setup	60
Figure 60: Analog Output Setup Menu	60
Figure 61: Analog Output Setup Screen	61
Figure 62: Graph Setup Menu	62
Figure 63: Graph Setup Screen	62
Figure 64: Scaler Editor Menu	63
Figure 65: GSF Warning	63
Figure 66: Scalar Editor Screen	64
Figure 67: Scheduled Reboot Menu Selection	64
Figure 68: Next Scheduled Reboot	65
Figure 69: Diagnostics Menu Selection	66
Figure 70: Active Zero Setup Screen	66
Figure 71: Test Relays Menu Selection	68
Figure 72: Test Alarm Relays Screen	69
Figure 73: Test Analog Outputs Menu Selection	70
Figure 74: Test Analog Voltage Output Screen	71
Figure 75: Signal Monitor Menu Selection	71
Figure 76: Signal Monitor Screen	72
Figure 77: Cell Pressure Calibration Menu Selection	73
Figure 78: Cell Pressure Calibration Screen	74
Figure 79: Manual Cell Pressure Override Menu Selection	75
Figure 80: Manual Cell Pressure Override	76
Figure 81: Power Up Default Menu Selection	77
Figure 82: Power Up Default Screen	78
Figure 83: Date/Time Menu	78
Figure 84: Date/Time Setup Screen	79
Figure 85: Date/Time Edit Screen	80
Figure 86: Communications Menu Selection	80
Figure 87: Communications Setup Screen	81
Figure 88: Download System Data Menu	81
Figure 89: System Data Download	82
Figure 90: Media Warning	82
Figure 91: Not ready Drive d:\	83
Figure 92: Download Timeline	83
Figure 93: System Info Menu	84
Figure 94: System Info Screen	85

Contents

Figure 95: Firmware Upgrade..... 85

Figure 96 The WEEE label 90

Figure 97 Schematic of Moisture Cell 91

Figure 98 Optional Hydrogen System Safety System..... 102

About this manual

Safety information

The following icons are used throughout this manual to identify any potential hazards that could cause serious injury to people or damage to the equipment:



This symbol warns of specific hazards which, if not considered, may result in personal injury or death.



This symbol warns of specific hazards from laser radiation.



This symbol warns of specific hazards from the presence of electrical voltage and current. .

Other information provided by the manual



This symbol highlights where you must take special care to ensure the Analyser or other equipment, or property is not damaged.

Note: *Notes give extra information about the equipment.*

Scope of the manual

This manual is the Installation and Operation Manual for the NANO Trace DF-745SGMAX analyzer. The DF-745SGMAX is also supported by the DF-700 Series Quick Start Guide, p/n NANOTRACE700Q.

Getting Started

Thank you for selecting the NanoTrace DF-745SGMAX Analyzer. Servomex designs, manufactures, exhaustively tests, and supports every analyzer under the tightest quality controls. You can expect every Servomex analyzer to arrive in perfect condition and provide years of trouble-free service if maintained properly. Please call your nearest Servomex Business Center if you need assistance.

Quality

The NanoTrace DF-745SGMAX Analyzer has been designed and manufactured under ISO-9001 controls, thus helping to ensure the highest possible standards of quality. Every analyzer that Servomex manufactures is tested to ensure that it functions properly when you receive it.

The optional certificate of calibration documents that your analyzer has been calibrated on gases that are traceable to NIST standards. With proper maintenance, your analyzer will remain calibrated for years.

For more information, or if you have questions, please do not hesitate to go to our website at Servomex.com, or contact your local Servomex Business Center as found on the back cover of this manual.

Usage

The NanoTrace DF-745SGMAX Analyzer is designed to measure the moisture content of a sample gas. This manual is written with the assumption that the user will be using the analyzer in this fashion. It is assumed that NanoTrace Analyzer users are familiar with the techniques and precautions associated with Ultra-High Purity (UHP) gas, its plumbing, and devices such as UHP regulators and gas purifiers, and that the analyzer is used as designed and intended.

For a fast and successful start-up, read this manual carefully. The manual contains important cautions, instructions, and guidelines that ensure the successful operation of your analyzer.

Installation and Maintenance

The NanoTrace Analyzer will provide years of accurate and dependable service if it is set up, operated, and maintained properly. It is essential to make a careful and complete installation as outlined in **Chapter 3, Installation, Startup and Shut Down**.

Unlike much UHP analytical equipment, the NanoTrace Analyzer does not require periodic maintenance for its moisture cell or other electronic components.

1. Safety Summary

1.1 General Warnings and Cautions

WARNING: HAZARDOUS AC



Hazardous AC voltages are present within this instrument. Do not remove the top or left side covers — no user-serviceable parts are located under the covers. Do not connect this analyzer to AC power until all signal cables and plumbing have been connected. Disconnect the analyzer from AC power before installing or removing external connections, the sensor, the electronics, or when charging or draining electrolyte. Disconnect the analyzer from AC power before moving the analyzer.

WARNING: AC POWER GROUND REQUIRED



Ensure that power cord ground wire is properly installed in connector and that connector is plugged into to a grounded outlet. Failure to provide AC power ground is a hazard that may result in electrical shock or death.

WARNING: CHECK AC OPERATING VOLTAGE



The analyzer has been factory set to operate on either 110 or 220 VAC. Check the voltage setting marked near the AC inlet. The analyzer may be damaged if operated on a different voltage than marked. Contact Servomex if you have received an analyzer that is set to operate on a different voltage your facility supply.

WARNING: ENSURE PROPER CARRYING, LIFTING AND MOUNTING



The weight of this analyzer is greater than 30 kg. Proper methods of lifting and carrying can help to protect against injury. Two individuals should lift and carry the instrument. Make sure to center your balance and properly space your feet a shoulder-width apart when lifting the instrument. Bend at the knees and make sure your back is straight. Grip the instrument with your fingers and palms and do not lift unless your back is straight. Lift up with your legs, not your back. Carry the instrument close to your body. Lower the instrument by bending your knees and keep your back straight.

The DF-745 is designed to be rack mounted. Ensure that the rack slides, or support rails, are rated to support a 30 kg analyzer. Use screws in all four front panel rack mount locations.

DANGER



Potentially hazardous AC voltages are present within this instrument. Leave all servicing to qualified personnel. Disconnect the AC power source when installing or removing external connections, the sensor, or electronic components.

CAUTION



Do not set up or operate the analyzer without a complete understanding of the instructions in this manual. Do not connect this analyzer to a power source until all signal and plumbing connections are made.

CAUTION



This analyzer must be operated in a manner consistent with its intended use and as specified in this manual.

CAUTION: FOR INDOOR USE ONLY



The DF-745SGMAX is for indoor use only. It should not be in an area where it will be subjected to particulate, condensed moisture, caustic atmosphere, temperature extremes, or operation outside of any limits listed in Chapter 8, Specifications.

1.2 Laser Safety



The DF-745SGMAX analyzer is a Class 1 laser device (per BS EN 60825-1:2014 and 21CFR 1040.10 and 1040.11 except for deviations pursuant to Laser Notice No. 56, 5 May 2019.)

It contains a laser emitter inside the moisture sensor, which is sealed and not user serviceable. Do not disassemble or modify the moisture sensor. Exposure to laser emissions may result in injury, including damage to the eyes.

1.3 Safety Labelling and General Marking

Location of Informational product marking and safety labelling on the DF-745SGMAX is shown in **Figure 1**, **Figure 2**, and **Figure 3**.




Figure 1: Analyzer Markings Shown in Rear Area View




Figure 2: Sensor Markings – Side




Figure 3: Sensor Marking – Front

2. Introduction

2.1 Product Description

The NanoTrace DF-745SGMAX Analyzer is an instrument designed for continuous UHP (Ultra High Purity) analysis of moisture content in process gases. It is housed in a steel chassis for rack-mounting. Its external features include a 6 x 9-inch LCD screen, push-button keys for user-interface, a gas panel containing ports for input and output of gasses (process and pneumatic), flowmeter (optional), connectors for digital and analog data output, and an IEC 60320 C-13 receptacle for connection to electrical power. The internal components include a moisture cell, valves (optional) and plumbing for routing of gasses, and electrical and electronic components. The moisture cell, which is a sealed, non-serviceable unit, contains a Class 3 laser for performing moisture analysis using absorption spectrometry. A description of the analytical techniques employed is provided in [Chapter 9 Theory of Operation](#).

2.1.1 Product Dimensions

The figures below (**Figure 1**, **Figure 2**, and **Figure 3**) provide dimensioned front, rear, and left side views of the DF-745SGMAX Analyzer.




Figure 4: DF-745SGMAX Analyzer (front)




Figure 5: DF-745SGMAX Analyzer (rear)




Figure 6: DF-745SGMAX Analyzer (left)

2.1.2 Internal Components

The major internal components of the DF-745SGMAX Analyzer are depicted in




Figure 7, below.




Figure 7: Major Internal Components

2.2 Specifications Summary

A summary of DF-745SGMAX characteristics is provided in the paragraphs below. Complete specifications are provided in **Chapter 10 Specifications**.

2.2.1 Moisture Detection

Lowest Detection Level (LDL): 5 ppb @ Constant Conditions

Display Resolution: .01 ppb

Accuracy: 0-100 ppm range: greater of: $\pm 5\%$ of reading or ± 5 ppb @ constant conditions.

Background Gas Compatibility: N₂, O₂, H₂, He, Ar, CO₂, CO, SF₆, Ne, Kr, Xe, N₂O, CF₄, C₂F₆, C₃F₈, C₄F₈, NF₃, and blends of up to eight of supported gases.

Electrical Requirements:

Standard: 100 to 120 VAC, 50/60 Hz, 5 Amps

Optional: 200 to 240 VAC, 50/60 Hz, 2.5 Amps

2.2.2 Laser

Max Power: 50mW/30mW

Beam Divergence: 436 mrad

Pulse Pattern: CW

Analyzer Product Classification: -Class 1 (per BS EN 60825-1:2014 and 21CFR 1040.10 and 1040.11 except for deviations pursuant to Laser Notice No. 56, 5 May 2019)

2.2.3 Environmental Operating Conditions

General: Indoor use only in a clean, dry, non-hazardous area of pollution degree 2, as defined in EN 60664-1

Altitude: 2000 meters (6500 feet), max

Ambient Operating Temperature: 10° C to 40° C (50° F to 105° F)

Humidity: 80% humidity for temperatures up to 31°C, decreasing linearly to 50% relative humidity at 40°C

3. Installation, Start-up, and Shutdown

3.1 Unpacking Procedure

Follow the procedure below to unpack your DF-745SGMAX NanoTrace Analyzer:

1. Examine the condition of the packaging and its contents. If any damage is apparent, immediately notify the carrier and Servomex. Do not proceed with the installation.
2. Check the contents against the packing slip to make sure the shipment is complete. Unattached equipment may be shipped with the analyzer in supplemental packaging. Shortages should be reported to Servomex immediately.
3. All standard DF-745SGMAX analyzers are shipped with the following:

Item	Part Number
Power cord ¹	2527-0314, 115 VAC Connector (US) 2527-0321, 220 VAC Connector (UK) 2527-0338, 220 VAC Connector (EU)
USB, Clip Flash Drive, 4 GB	220562
Gasket, VCR	210431
Instruction Manual (provided digitally on USB, Clip Flash Drive P/N 220562)	225486
Quick Start Guide (In digital format and stored on USB, Clip Flash Drive P/N 220562)	NANOTRACE700Q
Notes: ¹ Customer may select from US, UK, and EU style power cable when ordering. If other type is required, cable is customer supplied.	

4. Open the analyzer door, remove any shipping materials, and verify that nothing has come loose during transit.

CAUTION: CHECK AC OPERATING VOLTAGE



The analyzer has been factory set to operate on either 110 or 220 VAC. Check the voltage setting marked near the AC inlet. The analyzer may be damaged if operated on a different voltage than marked. Changing the operating voltage requires the analyzer to be returned to the factory.

5. Examine the voltage indicator on the rear panel to verify that the voltage is set as ordered.
6. Save the original container in the event you may need to ship the analyzer to another location or back to the factory (see **Chapter 6 Service**).

3.2 Analyzer Installation Overview

Installation of the analyzer requires the following general steps:

1. Mounting the analyzer.
2. Connecting the sample gas line to the analyzer inlet fitting.
3. Connecting sample gas outlet and by-pass outlet (optional) to facility vent.
4. Connecting the pneumatic pressure service to the pneumatic inlet fitting if required.
5. Wiring connections to data and relay outputs.
6. Making the power connection to the analyzer.

3.3 Analyzer Installation

The following steps describe the procedure for installation and initial startup of the analyzer. Installation may be performed only by qualified personnel.




The DF-745SGMAX is for indoor use only. It should not be in an area where it will be subjected to particulate, condensed moisture, caustic atmosphere, temperature extremes, or operation outside of any limits listed in **Chapter 8 Specifications**.

1. Position or mount the analyzer in working location (in instrument rack or on shop table) within 8 feet of gas sources.

NOTE



Use backup wrenches as shown in **Figure 8** when removing or installing VCR connectors, otherwise torque may be applied to analyzer plumbing and cause damage.



Backup
Wrench

Figure 8: Backup Wrench Use

2. If the analyzer is equipped with the optional Hydrogen Safety System (HSS), install the shut-off valve for the Hydrogen Safety System on the analyzer process gas inlet fitting. Use VCR gasket at

connection point. Refer to **Figure 10** for analyzer gas connection locations.

NOTE



VCR gaskets are one-time use only. Replace every time plumbing components are separated. Be sure gasket is clean and free of debris. Refer to **Figure 9**.



Figure 9: VCR Gasket (with supplied mounting bracket)

3. On HSS-equipped system, connect the stainless-steel sample gas feed to the open end of shutoff valve (not supplied) using VCR gasket. Please note that shutoff valve requires pneumatic pressure to open and allow gas to flow through it.




Figure 10: DF-745SGMAX Analyzer Gas Panel

4. If the analyzer is not equipped with the optional HSS, connect the stainless-steel sample gas feed to the process gas inlet fitting using VCR gasket and connect moisture outlet to facility vent with propylene tubing.
5. If the analyzer has the valves option, connect the pneumatic supply to the analyzer Pneumatic Gas Inlet.

NOTE



If the Hydrogen Safety System (HSS) is included, the sample outlet line must be made of stainless steel. See **Appendix B — Hydrogen Safety System** for more information.

6. If analyzer has the valves option, connect by-pass outlet to facility vent. If system has HSS, by-pass outlet tubing must be stainless-steel.
7. Connect analyzer to electrical power. Be sure to verify that you are connecting to the correct voltage

(110VAC or 220VAC) for your analyzer. Contact Servomex if you received an analyzer that is not compatible with your facility power.



Customer may select either US, UK, or EU type power cord at time of order. If other type is required, cable must be customer supplied. Customer-supplied cord must be rated for 250 VAC, 10 A with appropriate plug for location and an IEC 60320 C-13 receptacle for connection to the analyzer.

8. Turn on analyzer. After booting up, the analyzer goes into warm-up.

NOTE



Analyzers with isolation valves option are isolated during warm-up.

9. If analyzer has valves option, verify that sample gas is flowing through the bypass flow meter. If analyzer does not have valves option, verify that main screen indicates flow.
10. Allow the sample inlet to purge for at least 10 minutes.

11. If analyzer has valves option, establish sample gas flow by pressing the menu key and selecting **Restore Sample Gas Flow** as shown in **Figure 11**. Verify that the main screen indicates flow.




Figure 11: Starting Gas Flow

3.4 Analyzer Start-up

Open the door and turn on the power using the main power switch inside the analyzer (see **Figure 7**). The analyzer will undergo a series of Diagnostic Procedures while the various start-up screens are displayed. Next, the Servomex logo is briefly displayed and then the data displays appear with the **Warming Up** message flashing (see **Figure 12**). The warm-up process takes approximately 6 to 15 minutes. After warm-up is complete, the display is like **Figure 18** (values will vary). The analyzer may take up to 90 minutes to completely warm up depending on environmental conditions.

NOTE:



During the warm-up period all analog and digital outputs are held to an artificial 0.011 ppb reading to avoid the reporting of false readings.




Figure 12: Data Display Screen

3.4.1 Gas Flow and Pressure Regulation

The gas delivery system is designed to operate with a gas flow rate of 1 liter per minute through the moisture sensor and with sample gas at ambient pressure (nominally 760 Torr).

3.4.2 Gas Flow Setting (without Isolation Valves)

For analyzers without isolation valves, set gas flow by adjusting a valve (customer-supplied) upstream of the analyzer while viewing the analyzer's internal flowmeter. Adjust flow to 1 SLPM.

3.4.3 Gas Flow Setting (with Isolation Valves Option)

NOTE: On units equipped with isolation valves, the state of the gas control valves for the moisture sensor is indicated on the main display. The default state of a factory-shipped instrument is isolation of the sensor. This can be adjusted as described in **4.2.13 Power Up Default**.

1. Restore sample gas flow to moisture sensor by selecting **Restore Sample Gas Flow** as shown in **Figure 11**.
2. Adjust the analyzer's regulator (see **Figure 10**) to achieve a flow rate of 1 SLPM using the internal flow meter as a reference.

On initial start-up, after the 6-minute warm-up process is complete, the moisture analyzer will take a few hours to equilibrate, and the concentration of moisture is automatically shown in parts per million (ppm) on the display.

If, after initial start-up, the power is briefly turned off, the analyzer will take approximately 6 minutes to re-equilibrate.

3.4.4 Back Flow Prevention System (Isolation Valves Option)

The Back Flow Prevention System (only on analyzers with isolation valves) provides protection against the possibility of contamination of the plumbing system when gas flow is lost or accidentally shut off. Under this condition, ambient air may be drawn into the system through the bypass exhaust by means of the aspirator or vacuum pump.

During normal operation, gas flow is constantly monitored by a flow switch mounted in the bypass loop that is set to trip when the flow drops below 0.1 lpm. When flow is reduced, the switch will automatically activate the gas panel isolation valves. In addition, the pressure reading will be forced to zero. When the system detects a zero-pressure reading, a message will be displayed which states that backflow prevention is engaged and to check sample and bypass flow. See **Figure 13**. The event also triggers a system alarm (192) for the moisture cell. See page 53 for additional information on system alarms.

The analyzer will remain in isolation and the user will not have control of any valves at this point.

CAUTION



While the Back Flow Prevention System is engaged the moisture reading is not valid.




Figure 13: Back Flow Prevention Warning

The user must determine the cause of the loss of gas flow and remedy the situation. When flow has been restored to the analyzer, the pressure reading will go to a non-zero value and a message will then appear stating that sample flow has been restored and to wait five minutes to restore internal sample flow to the sensors. See **Figure 14** .

A five-minute timeline is displayed, and the user will still not have valve control until the five minutes are up. During this time, the **Esc** key may be hit to cancel the five-minute delay and return control to the user. Once the Esc key is hit, or the five-minute waiting period has expired, all remaining messages are cleared from the screen and the system will remain in isolation until the user restores flow. See **6.3.5 Restore Sample Gas Flow** for additional information on restoring gas flow.




Figure 14: Re-established flow delay

3.4.5 Download System Data

The final step of the installation, after a couple days of dry down, is to download the system data files (see **6.3.19 Download System Data**) which may optionally be sent by e-mail to Servomex for review. This will allow the factory to confirm that the analyzer is working properly by comparison with data stored at the time of shipment, and in addition will set a baseline for comparison with future downloads, if any.

3.5 Analyzer Shut Down

To minimize the time required for the analyzer to re-achieve a zero baseline on start-up, the following steps should be followed when shutting the analyzer down.

Short Term Shut Down. A short-term shut down, for example to move and restart the analyzer, can be accomplished by simply shutting off the power switch behind the front door. This action initiates the standard isolation process (for analyzers equipped with gas control valves) and computer shut down which takes approximately 40 seconds.

Long Term Shut Down. For analyzers with gas control valves, from the Main Menu go to System, and select Isolate Analyzer. A routine automatically starts that closes the downstream valve and allows pressure to build in the moisture cell after which an upstream valve closes as well. Throughout this process a message appears over the display instructing the user to wait 40 seconds. Once complete, the message disappears, and the user may shut the power off with the switch behind the front door that initiates the computer shut down sequence that takes an additional 40 seconds.

Additional long-term isolation security can be achieved by closing the valve on the rotameter on the rear of the instrument as well as completely closing the sample inlet regulator.

If the analyzer is being disconnected from gas, be sure to tightly cap all gas connections. Refer to **Figure 16** for a reference diagram of the gas flow path including isolation valves.

NOTE: See the section on Power up Default on page 73 for \setting user selectable preferences at the time of power up.

Loss of power will result in automatic valve closure and the following restoration of power may result in the "Scan Disk" function occurring before system start-up.




Figure 15: Block Diagram of Gas Flow Path




Figure 16: Block Diagram of Gas Flow Path with Optional Valves

4. Options

The following options for the DF-745SGMAX are available at the time of order:

4.1 Key Lock

An optional key lock can be installed in the door of the analyzer to prevent access to the power switch and other internal components. The lock is supplied with two keys.

The key lock only prevents access to the internals of the analyzer. It does not prevent adjustments through the interface on the front panel.

4.2 Operating Voltage

The analyzer can be wired for operation at either 100-120 Volts AC or 200-240 Volts AC.

The operating voltage is not adjustable in the field.

4.3 Serial Communications

The analyzer can be set to communicate through either RS-232 or RS-485 protocols.

The serial communications option is not adjustable in the field.

4.4 Analog Voltage Output

The analog voltage output can be set at the factory to 0-1, 0-5-, or 0-10-volts DC.

The analog voltage output is not adjustable in the field.

4.5 Isolation Valves

The isolation valves option allows the moisture sensor cell and internal gas lines to be isolated (physically separated) when the inlet or outlet of the analyzer is exposed to ambient air so that they will not be contaminated with moisture. This is accomplished with a system of valves that are operated by selecting **Isolate Analyzer** from the main menu. A block diagram showing gas flow in a system with the isolation valves option is provided in **Figure 16**.

4.6 Hydrogen Safety System

This option enables the analyzer to be safely used in a hydrogen background application. See **Appendix A – Hydrogen Safety System** for additional information.

5. Connecting to External Devices

The analyzer can be interfaced to a variety of external devices via the ports on the rear panel. Alarm contacts, voltage, and current outputs, and serial communications are supported. All outputs, analog or digital, are fully isolated from earth ground. Refer to **Figure 17**.

NOTE



During the six-minute warm up period at startup, all analog and digital outputs are held to an artificial 0.011ppb reading to avoid the reporting of false readings.




Figure 17: Rear Panel Electrical Connectors

5.1 Serial Communication Port – J5

Either of two serial communication ports are available at the time of order: RS232C or RS485 which enable interface between the analyzer and other operating systems.

Up to 32 units may be accessed via RS-485. Operating parameters are 8 bits, no parity, and one stop bit. Baud rate may be selected from the menu on the display.

See **Table 2** on page 25 for wiring information.

See the chapter on Communications on page 78 for additional information on setting unit ID's and baud rates. A program to facilitate serial communications is available from Servomex. Please contact us the applicable email for your region: service_americas@servomex.com (North and South America), service_europe@servomex.com (Europe), or service_asia@servomex.com (Asia).

Pin #	Signal	Description
J5-8	GND	Ground
J5-7	TX-	4 wired 485 paired with TX+
J5-6	Key	Unused
J5-5	RX-	4 wire 485 paired with RX+
J5-4		Unused
J5-3		Unused
J5-2	TX+	Data transmitted by the analyzer via RS-232 or RS-485
J5-1	RX+	Data received by the analyzer via RS-232 or RS-485

Table 1: Pin-out of Serial Comm Connector J5

Pin assignments			
DF-745SGMAX Connector J5	PC-DB9/RS-232	PC-DB25/RS-232	PC/RS-485 converter
1: RX+	3: TD	2: TD	TX+
2: TX+	2: RD	3: RD	RX+
8: Gnd	5: Gnd	7: Gnd	
5: RX-			TX-
7: TX-			RX-

Table 2: Serial Communications Connections

5.2 Analog Signal Outputs – J4

The analog voltage output correlating to the front panel display reading is provided on the rear of the analyzer through connector J4. The full-scale voltage is set at the factory at the time of order to: 0 to 1 VDC, 0 to 5 VDC, or 0 to 10 VDC. The output is electrically isolated from all other analyzer outputs, and from chassis (Earth) ground.

5.3 4-20 mA Outputs – J4

J4 Pin #	Moisture Signal	Description
J4-8	GND	Ground
J4-7	Key	
J4-6	A Out +	Analog Voltage Output (+)
J4-5	A Out -	Analog Voltage Output (-)
J4-4		Unused
J4-3		Unused
J4-2	4-20 mA +	4-20 mA Output (+)
J4-1	4-20 mA -	4-20 mA Output (-)

Table 3: Pin-Out of Moisture Signal Output Connector J4

The 4-20 mA analog output correlating to the front panel display reading is provided on the rear of the analyzer through connector J4. This output is electrically isolated from all other analyzer outputs, and from chassis (Earth) ground.

The maximum load resistance for each is 1K Ohms and the analyzer provides a compliance voltage of approximately 28 VDC.

5.4 Relay Ports – J8, J9

Four form C (SPDT) relays (contact closures) are provided to assign to the various alarms. The contacts are rated at 30 VDC, 1A. They are not designed to switch AC power.

The relay contacts can be programmed through the user interface for up to four moisture levels, temperature, moisture cell diagnostics, zero calibration in progress, analyzer off line, or freeze of analog outputs during calibration.

The relays are wired for Fail Safe operation such that a Normally Open (No alarm) contact connects to common when an alarm occurs or when power to the instrument is lost.

The relay wiring can be tested with the Relay test routine found on page 68.

Pin #	Moisture Relay	Description
J8-8	GND	Ground
J8-7	RLY2-NC	Relay 2 Normally Closed
J8-6	RLY2-NO	Relay 2 Normally Open
J8-5	RLY2-COM	Relay 2 Common
J8-4	RLY1-NC	Relay 1 Normally Closed
J8-3	RLY1-NO	Relay 1 Normally Open
J8-2	RLY1-COM	Relay 1 Common
J8-1	KEY	Unused
J9-8	GND	Ground
J9-7	RLY4-NC	Relay 4 Normally Closed
J9-6	RLY4-NO	Relay 4 Normally Open
J9-5	RLY4- COM	Relay 4 Common
J9-4	RLY3-NC	Relay 3 Normally Closed
J9-3	RLY3-NO	Relay 3 Normally Open
J9-2	Key	Unused
J9-1	RLY3-COM	Relay 3 Common

6. User Interface

6.1 Data Display Screen

The 10.1 in. LED display on front panel provides output of information to the user. The display format during normal operation of the analyzer is shown in **Figure 18** below.



Figure 18: Data Display Screen

Display elements include:

- Alarm status indicator – **OK** denotes no alarm conditions. Message and error code displayed in event of alarm.
- Most recent H₂O measurement (large numerals).
- A strip chart history of concentration measurements.
- A concentration output range for the analog outputs, as designated by the user.
- Background gas as set by user.
- Gas pressure within the moisture sensor

The digital readout of the moisture concentration will be overwritten with a warning if any of the level or system alarms are tripped.

6.2 Keypad

The keypad allows the user to control all the features of the analyzer. The layout of the keypad on the front panel is represented in **Figure 19** below.




Figure 19: Keypad

The **Menu** key activates the menu structure along the top of the GUI interface on the view screen.

Once in the menu, the arrow keys (▲ and ▼) highlight the various menu features. When the desired selection is highlighted, the right arrow will access the submenu if available (denoted by a right arrow next to the menu text). The '**Next**' key and the '**Enter**' key will do this as well. The arrows also enable the entry of numerical parameters.

The **Enter** key will call up dialogue boxes from the menu (denoted by the sequence "... "next to the menu text). It will also enter numerical values within dialogue boxes.

The **Next** key allows the user to change between active inputs within a dialogue box.

The **ESC** key allows the user to exit numerical entry boxes within dialogue boxes without any user changes, dialogue boxes without any user changes, and the menu bar.

6.3 Menu Structure

The NanoTrace Moisture Analyzer menu tree consists of main menus, sub-menus and screens and is depicted in **Figure 20** below.

6.3.1 Main Menu




Figure 20: Main Menu (Without Isolation Valves Option)

The Main Menu is accessed by pressing the Menu key on the front panel. Use the arrow keys (▲ and ▼) to navigate up and down through the list. Select the highlighted item with the **Enter** key on the front panel

Figure 20 shows the main menu for the standard DF-745SGMAX, which does not have isolation valves installed. The **Isolate Analyzer**, **Restore Sample Gas Flow**, and **Power Up Default** selections are greyed-out because these features are not available on models without isolation valves.

The Main Menu for a DF-745SGMAX with the Isolation Valves Option is shown in **Figure 21**, below. **Isolate Analyzer**, **Restore Sample Gas Flow**, and **Power Up Default** are available on this model.




Figure 21: Main Menu (With Isolation Valves Option)

6.3.2 Set Up GSF

The **Set Up GSF** menu selection is shown in **Figure 22**, below. This menu selection is used to access the Sample GSF Setup Screen as shown in **Figure 23**.




Figure 22: Setup GSF Menu

The Gas Scale Factor (GSF) editor is used to define a new gas mixture for analysis or to access a predefined mixture template for analysis.

Defining a gas mixture for moisture analysis can be accomplished in two ways. The user can select gases from a table of thirteen supported gases, and then specify the % volume for each gas, or the user can select a pre-defined mixture called a template. Both options are available through the GSF Editor Screen by means of the 'Supported Gases' display as shown in Figure 23 and the **Template** display as shown in Figure 24.

To switch between the **Supported Gases** display, and the **Template** display, use the **Up-arrow** key to move to the top of the **Supported Gas** or **Template** list, at which point the arrow on the screen above the list will turn red. Press the **Up-arrow** key one more time and the screens will toggle between **Supported Gases** and **Template** mode. Note that the red arrow key in the upper left part of the screen is displayed only when the top item is selected in the **Supported Gases** or **Template** screen.

To navigate to different areas on a screen (tables and buttons), use the **Next** key on the front keypad.




Figure 23: Sample GSF Setup Screen

6.3.2.1 Supported Gases Display – Specifying a Mixture

The Supported Gases display in the GSF editor, as shown in **Figure 23**, is used to specify a new mixture for analysis by first identifying the gases in the mixture and then identifying the percentage of each gas in the mixture. Note - pressing **Esc** at any time, will return the user to the Data Display screen canceling any changes made to the GSF Editor to that point.

- **To identify gases:**
 1. Select (highlight) each gas to be included in the mixture from the "Supported Gases" table using the **up** and **down Arrow** keys (▲ and ▼). *
 2. Use the **right arrow** key to add a highlighted gas to the "Sample Gas Mixture" table. **Note:** - a maximum of 8 gases can be specified in a mixture.
 3. Use the **left arrow** key to delete highlighted gases from the "Sample Gas Mixture" table. Note – if no gases are highlighted, then the bottom-most gas in the mixture table will be deleted every time the **left arrow** key is pressed.
 4. When the "Sample Gas Mixture" table is complete, accept the mixture by using the **Next** key to navigate to the Accept Mixture button and press **Enter**.
- **To set the volume of each gas:**
 1. Use the **Left** and **Right** arrow keys to select the digit to be changed.
 2. Use the **Up** and **Down** arrow keys to change the value of the selected digit.
 3. Select the next gas to be edited by using the **Next** key and follow the same procedure as above to set the percentage for each gas.
 4. When all the % values have been entered, use the **Next** key to navigate to the Accept button and press **Enter**. Note - the Accept button is activated only when the gas % sum is between 99.9 and 100.1%. If the sum is outside of this range, the sum is shown in red and the Accept button is deactivated and shown as light gray.
 5. Balance Update feature - Any gas % left as 0 can be automatically set by the analyzer using the Balance Update feature as a convenience to the user.

- a. -For a single gas, the Balance Update will calculate the % required to make the sum of all gases equal to 100%.
 - b. -If the percentage of two or more gases are not specified, the Balance Update feature will calculate equal values for each gas to bring the sum of all gases to 100%.
 - c. When the percentage of each gas in the mixture has been properly set, use the **Next** key to navigate to the Accept button and press **Enter** to save the mixture parameters and return to the main data display.
6. The GSF setup is then saved upon acceptance into a tab-delimited line in the monthly GSF log file as described on page 50.
- *There are 13 standard gases in the Supported Gases table. Up to 10 additional user-specified gases can be added to the table. Please contact Servomex Support for further information.

6.3.2.2 Template Display – Creating/Applying/Editing/Deleting a User Defined Mixture

The **Template** display in the GSF editor, as shown in Figure 24, enables the user to create, edit or recall pre-defined templates of gas compositions for easy retrieval. Note - pressing **Esc** at any time will return the user to the Data Display screen canceling any changes made to the GSF Editor to that point. Access to the template display screen is gained through the Support Gases display. To switch between the **Supported Gases** display, and the **Template** display use the **Up-arrow** key to move to the top of the supported gases list, at which point the arrow on the screen will turn red. Press the **Up-arrow** key one more time and the screens will toggle between **Supported Gases** and **Template**. Note that the red arrow key in the upper left part of the screen is displayed only when the top item is selected in the **Supported Gases** or **Templates** screen.




Figure 24: GSF Editor with Template Active

6.3.2.3 Creating a new template

1. To create a new mixture template, first define the gases by following steps 1-4 in section 2057342997. .mmmmmmmmmmcmxlxvii.3. and then define the percentages of each gas by following steps 1-4 in section 2057342997. .mmmmmmmmmmcmxlxvii.3.

2. Use the **Next** key to navigate to Balance Update button and hit Enter.
3. Name the new template by activating the soft keyboard by hitting the up arrow when:
 - a. The arrow to the left of the Balance Update button is red
 - b. The Balance Update button is blinking
 - c. The sum of the gas percentages equal 100% (within 99.9 and 100.1)
 - d. Fewer than 30 templates have been saved

6.3.2.4 Naming a Template

A template name must have a minimum of 3 and a maximum of 20 characters.

1. Create a name for the new template by using the **Arrow** keys to select each character and use the **Next** key to accept or clear the selected character.
2. The display shows the name entry and the number of characters. They are in green when between 3 and 20, otherwise they are in orange.
3. Accept the entry as a new template name. If the template name already exists, it overwrites without confirmation. If the template name has fewer than 3 characters, no new template is generated.
4. **Limitation:** The soft keyboard will only be displayed for the first time after activating the "GSF Editor". That is, every visit to the "GSF Editor" allows saving one template only. If the soft keyboard is inadvertently activated after its 1st use, press the **Esc** key to get back to the Editor.

6.3.2.5 Applying an existing template

1. The GSF Template screen displays available user-defined templates (up to 30 are allowed). Use the **Up** and **Down** arrow keys to highlight a template from the list and to display the recipe (gases, percentages) in the table to the right of the template list. See section 6.3.2.3 for instructions on creating a new template.
2. Use the **Right arrow** key to select a particular template. Selecting a template populates the Sample Gas Mixture table in the Supported Gases display.
3. When the proper template has been selected, use the **Next** key to navigate to the Accept Mixture button and press **Enter**.
4. The template values (gases, %) can be used as defined or can be modified for use as a special mixture. The latter is accomplished using steps 1-7 in Section 7.
5. When the percentage of each gas in the mixture has been properly set, use the **Next** key to navigate to the Accept button and press **Enter** to save the mixture parameters and return to the main data display.
6. The GSF setup is then saved upon acceptance into a tab-delimited line in the monthly GSF log file as described on page 50.

6.3.2.6 Editing the makeup of an existing template

1. Access to the Template display screen is gained through the Support Gases display as described in section 6.3.2.1.
2. To switch between the 'Supported Gases' display, and the 'Template' display, use the **Up-arrow** key to move to the top of the supported gases list, at which point the arrow on the screen will turn red. Hit the **Up-arrow** key one more time and the screens will toggle between Supported Gases and Template. Note that the red arrow key is displayed only when the top item is selected in the "Supported Gases" or "Templates" screen.
3. Use the **Up** and **Down arrow** keys to highlight a template from the list.

4. Use the **Right arrow** key to select a particular template which populates the Sample Gas Mixture table in the Supported Gases display.
5. Edit the contents of the mixture as described in **6.3.2.1 Supported Gases Display – Specifying a Mixture** and in **6.3.2.2 Template Display – Creating/Applying/Editing/Deleting a User Defined Mixture**.
6. Rename the template if desired as described in **6.3.2.4 Naming a Template**.
7. To use the edited template, but not rename it, navigate with the **Next** key to the **Accept** button and press enter.

6.3.2.7 Deleting an existing template

- 1) The GSF Template screen displays available user-defined templates (up to 30 are allowed). Use the **Up** and **Down arrow** keys to highlight a template from the list and to display the recipe (gases, percentages) in the table to the right of the template list.
- 2) Use the **Left arrow** to delete a template from the table. Deletion of a template prompts a confirmation.

After the percentages of all background gas are entered, the **Accept** button is hit and the system confirms that the total is 100%. See **6.3.10 Analyzer Setup** for additional information on setting the Pressure Alarm.

NOTE: An entry of any percentage of Hydrogen in the GSF calculation will automatically engage the Hydrogen Safety Service System option if equipped. See **Appendix A – Hydrogen Safety System** for additional information.

6.3.3 HFC Samples

The HFC Samples Menu selection is shown in **Figure 25**, below. This menu selection accesses the Interfering Sample Gas selection dialog box shown in **Figure 26**.




Figure 25: HFC Samples Menu Selection




Figure 26: Interfering Sample Gas Selection

This dialog box allows the user to select a background of one of 3 HFC gases, CHF3, CH2F2, or CH3F. These gases are differentiated from the main GSF menu due to the spectral interferences from the HFC's as well as varying levels of moisture that are commonly present with them. As these are typically stable in any given batch of product, their contribution can be removed with an offset. Select the applicable HFC gas and then select **Accept** and press **Enter**. The **Baseline Status Action** dialog box will be displayed as show in **Figure 27**.




Figure 27: Baseline Status Action Dialog Box

If you have not yet measured moisture in the selected CHF gas, you will need to gather the baseline data. This is done by putting a purifier in-line to remove all moisture, then selecting **Gather CH-- baseline**. Servomex offers an externally mounted gas panel with a purifier that can be used for this purpose (Servomex P/N 212896). After installing purifier, select **Gather CH-- baseline**, then **accept**, and press **Enter** to continue. The **Baseline Data Screen** will be displayed, as shown in **Figure 28**.




Figure 28: Baseline Data Screen

Select **Accept** and press **Enter** to accept the baseline manually, or allow analyzer to run until it automatically accepts. After acceptance, uninstall the purifier and proceed to analysis.

6.3.4 Isolate Analyzer (Isolation Valves Option)

The **Isolate Analyzer** menu selection is shown in **Figure 29**, below.




Figure 29: Isolate Analyzer

The **Isolate Analyzer** feature is available only on analyzers with the isolation valves option. This function allows the internal gas lines and the moisture sensing cell volume to be isolated for a planned removal of the external delivery lines. It is also considered the first step in the shut-down procedure.

Select **Isolate Analyzer** and hit the **Enter** key to initiate the isolation process. The Isolate Warning shown in **Figure 30** will appear for 40 seconds.




Figure 30: Isolate Warning

When the moisture cell is isolated from gas flow, the **Isolated** message appears at the bottom of the main display and the pressure alarm is disabled.



When in isolation mode, the moisture cell's pressure transducer will yield an incorrect reading of the actual cell pressure. Because the displayed moisture reading is dependent upon a proper pressure reading, moisture readings should be ignored while the cell is under isolation.

NOTE: Under these conditions the pressure alarm is disabled.

6.3.4.1 Back Flow Prevention System (analyzers with isolation option)

The Back Flow Prevention System provides protection against contamination of the plumbing system when gas flow is lost or accidentally shut off. Under these conditions, ambient air may be drawn into the system through the bypass exhaust by means of the aspirator.

During normal operation, the gas flow is constantly monitored by a flow switch mounted in the bypass loop that is set to trip when the flow drops below .1 lpm. When flow is reduced, the switch will automatically activate the gas panel isolation valves. In addition, the pressure reading will be forced to zero. When the system detects a zero-pressure reading, a message will be displayed which states that backflow prevention is engaged and to check sample and bypass flow. See **Figure 31**. The analyzer will remain in isolation and the user will not have control of any valves at this point.




Figure 31: Back Flow Prevention Warning

The user must determine the cause of the loss of gas flow and remedy the fault. When flow has been restored to the analyzer, the pressure reading will go to a non-zero value and a message will be then appear stating that sample flow has been restored and to wait five minutes to restore internal sample flow to the sensors. See **Figure 32**.

A five-minute timeline is displayed, and the user will still not have valve control until the five minutes are up. During this time, the **Esc** key may be hit to cancel the five-minute delay and return control to the user. Once the Esc key is hit, or the five-minute waiting period has expired, all remaining messages are cleared from the screen and the system will remain in isolation until the user restores flow. See **6.3.5 Restore Sample Gas Flow** for additional information on restoring gas flow.




Figure 32: Re-established flow delay

This event will trigger a system alarm (192) for the moisture cell. See page 53 for additional information on system alarms.

6.3.5 Restore Sample Gas Flow (Isolation Valves Option)

The **Restore Sample Gas Flow** menu selection is shown in **Figure 33**, below.




Figure 33: Restore Sample Gas Flow

Restore Sample Gas Flow is available only on analyzers with the isolation panel option. This selection allows the user to return the analyzer gas flow to normal after isolation.

6.3.6 Calibrate Menu

The **Calibrate** menu selection is shown in **Figure 34**, below.




Figure 34: Calibrate Menu

Instrumental drift has minimal effect on the quantitative moisture content result. Once operating parameters are in place from the factory, no other SPAN calibration is necessary.

Sample line contributions to moisture, however, especially at the sub-ppb level, are difficult to remove even in a correctly operating system. The dry down process of the analyzer and/or gas supply system can take many months. The moisture analyzer has an array of zeroing features that enable the user to establish temporary performance near 0.0 ppb during the dry down process. Monitoring the zero-reference number during that time, or repeated restoration of the factory zero, will give the user a sense of the offset being induced by the user zero actions.

Any zero action whether completed or aborted is recorded in the **Zero Cal Log** as shown in **Figure 44**.

6.3.6.1 Check/Adjust Zero

The **Check/Adjust Zero** menu selection is shown in **Figure 35**, below. This menu selection accesses the Check/Adjust Zero screen, shown in **Figure 36**.




Figure 35: Check/Adjust Zero Menu

The **Check/Adjust Zero** screen displays many pieces of information including a live reading of moisture in ppb (or ppm) and **Zero Reference** and **Zero Offset** values.




Figure 36: Check/Adjust Zero Screen

The **Zero Reference** value is based on the deviation of the instrument's baseline from what was set at the factory. The zero reference of an instrument from the factory will be 0.00. After a manual or auto zero is performed, this value may change slightly.

If the moisture cell has been previously calibrated by the user, **Restore Factory Zero** will be enabled on this screen. Enacting this will erase any user zero and restore the factory set point.

The **User Zero Offset** and **Active Zero Offset** are two mechanisms to modify the zero baseline. When the Active Zero feature is turned on, the **User Zero Offset** display is inactive and **Active Zero Offset** will be displayed instead. Refer to **4.2.12.1 Active Zero On/Off** for additional information on the **Active Zero** feature.

A relay is available on the Analog Output Setup Screen (see page 61) to signal that a zero calibration is taking place, and the analog output signal can also be frozen or allowed to update during the calibration process.

6.3.6.1.1 User Zero Offset

The **User Zero Offset** function enables the user to add a given moisture ppb value to the displayed concentration. This can be useful in preventing inaccurate readings to be displayed should the baseline drift below the calibrated zero point. The value does not affect zero calibration, it is simply added to the calibrated zero. For example, an offset of 1.0 ppb could be put into a system reading 0.0 ppb to allow a chart recorder attached to the output to read slightly above zero. Under this condition, the moisture reading would be 1.0 ppb.

Use the arrow keys (▲ and ▼) to highlight **Enter User Zero Offset** in the **Check/Adjust Zero** screen (**Figure 36**). Then press the **Enter** key on the front panel, and the zero offset box will appear on the screen. With the left and right arrow keys, move the cursor to the right of the digit you want to change. With the up and down arrow keys set the number to the desired value. When done, hit the **Enter** key which will move the highlighted area to the **Apply** button and hit **Enter** to set the value. Use the **Next** key to go back and change the value or move to the **Done** button, followed by pressing the **Enter** key to leave the screen. Using the **ESC** key at any time will exit the screen making no changes and return to the main display.

6.3.6.1.2 Do A Manual Zero

The manual zero command enables the user to zero the moisture cell reading by applying a correction to the analyzer measurement.




Figure 37: Manual Zero Screen

For this purpose, it is necessary to ensure that moisture free gas (sub ppb) is entering the sensor through the process inlet. When switching to a gas that is moisture free, it is important to wait the necessary time to allow the reading to re-stabilize.

It is extremely important to note that moisture is inherently slow to move and as a result the process of doing a zero calibration can take many hours before the reading is at baseline. A premature zero, although acceptable, will ultimately result in an inaccurate reading as the plumbing and analyzer continue to dry out requiring another zero action.

After selecting **Do a Manual Zero**, a screen will appear which displays a trace of the recent moisture reading. See **Figure 37**. Observe the trace until the reading is stable and then press **Enter**. This action will accept the present value as the new zero setting for the moisture cell and the **Zero Reference** field will be updated. After this action, the user will be brought back to the **Check/Adjust Zero** screen.

Pressing **ESC** during the calibration process will abort the action and return to the main display. Whether the calibration is complete or aborted, pressing **Done** from the previous screen will allow the user to return to the main display.

Once complete, the gas source must be returned to the original state with process gas going directly through the sensor.

A relay is available on the Analog Output Setup screen (see page 61) to signal that a zero calibration is taking place, and the analog output signal can also be frozen or allowed to update during the calibration process.

6.3.7 Data History Routine

The **Data History** menu selection is shown in **Figure 38**, below.

The Data History Screen (**Figure 39**) enables the user to see the data history displayed in strip chart form on the front display. By default, the data history screen displays data for the most recent 24-hour period sampled at 1 point per minute (fixed) and the y-axis is auto-scaling.

The **Next** button can be used to toggle the X axis from 1 day (default), to 1 hour, to 1 week and then to 3 weeks.

The **Next** button can also toggle the cursor to the max and min values on the Y axis and the arrow keys can then be used to adjust the values, and the display will actively update.




Figure 38: Data History Menu

The data history may be downloaded to a USB memory stick by using the **Next** key to move to the **Download** box and hitting **ENTER**. A screen will appear, requesting that a memory stick can be placed in the external USB socket. The socket is located behind the front door on the left side of the chassis.

The downloaded file will be in tab delimited form and will be all moisture data in the system up to 3 weeks old if available. The download process will take up to 15 seconds and the display will indicate downloading is in progress. Once the download is complete, control of the analyzer is returned to the operator.




Figure 39: Data History Screen




Figure 40: Install Media

Figure 41: Example of Data Download

6.3.8 Data Downloader Routine

The **Data Downloader** menu selection is shown in **Figure 42**, below.




Figure 42: Data Downloader Menu

When the user selects **Data Downloader** from the main menu, the Data Downloader screen, **Figure 43**, displays. This screen enables the user to label data with unique location names as well as to view and download specified data.

The **Next** key is used to toggle through the various options on the screen and the arrow keys (▼ and ▲) move up and down through the location list.




Figure 43: Data Downloader Screen

6.3.8.1 Set Location

The **Set Location** function is used to choose a location from a list of existing locations previously entered the system (see **6.3.8.3 Add Location**). On the Moisture Data Downloader screen, Figure 43, use the **Next** key to move to the list of existing names and then use the arrow keys (▲ and ▼) to select the location desired. Then, use the **Next** key to move to **Set Location** and press **Enter** to accept the new location.

The action of setting a location starts the logging process and creates a new file. Changing to a new location will, in turn, end the previous file and start a new one.

6.3.8.2 View Location




Figure 44: View Location Screen

The view location screen (**Figure 44**) is used to view data previously stored in the system sorted by location. On the Data Downloader screen (**Figure 43**) use the **Next** key to move to the list of existing names and then use the arrow keys (▲ and ▼) to select the location desired. Then use the **Next** key to move to **View Location** and press **Enter**. The View Location screen will appear as in **Figure 44**.

Use the arrow keys (▲ and ▼) to select the data block desired and use the **Next** key to move to **View Location** and press **Enter**. The data history screen will appear as shown in **Figure 39**. From the data history screen, the data may also be downloaded to a USB Memory Stick.

6.3.8.3 Add Location

The user can create a new location stamp by moving the cursor to the Add Location button and hitting enter. This brings up the keyboard shown in **Figure 45** that is used to enter the name of the new location.




Figure 45: Keyboard Display

Use the arrow keys (▲ and ▼) to navigate the keyboard and use the **Enter** key to accept each character. If an error is made, use the **Next** key to move to the **Clear Entry** key and hit enter. When the location name is complete, use the **Next** key to move the highlight to **Accept New Location** and hit **Enter**. The display will return to the Downloader Screen and the name will appear in the list of available locations.

6.3.8.4 Delete Location

The delete location function is used to remove a location from the list of available names. On the Data Downloader screen (Figure 43), use the **Next** key to move to the list of existing names and then use the arrow keys (▲ and ▼) to select the location desired. Then use the **Next** key to move to **Delete Location** and press **Enter**. A confirmation box will then appear (see Figure 46) and the user can either accept the deletion with the **Enter** key or can hit ESC to cancel the action. If accepted, the name will be removed from the list of available locations.




Figure 46: Delete Selection

6.3.9 View Logs Menu

The **View Logs** menu selection is shown in Figure 47, below.




Figure 47: View Logs Menu Selection

View Logs allows the user to easily access past events that may be related to past operational changes (e.g., zero) or instrument upsets. Use the arrow keys (▲ and ▼) to scroll up and down through the list. Pressing **ESC** will return to the main display.

6.3.9.1 Zero Log

The **View Zero Log** menu selection is shown in **Figure 48**, below.




Figure 48: View Zero Log Menu

The Zero Log (**Figure 49**) reports on adjustments made to the moisture cell zero setting. The date and time of the zero calibration is noted.




Figure 49: Zero Log Screen

6.3.9.2 View System Error Codes

The **View System Error Codes** menu selection is shown in **Figure 50**, below.




Figure 50: View System Error Code Log Menu




Figure 51: System Error Log Screen

The System Error Log (**Figure 51**) reports functional errors in the moisture system. If the error persists for more than 30 minutes, the code is displayed. In addition, a system alarm will trip if configured to do so. See page 71 for additional information on setting System Alarms. **NOTE:** The 30-minute clock is delayed for 60 minutes on a “cold” system start up.

Table 5 provides common error codes that may be reported on the **System Error Log**.

Error Number	Description
101	Moisture Peak Not Found
121	Low Sample DC
161	Laser Temperature too far from Setpoint
191	Fan condition
192	Backflow Prevention – ensure adequate flow through the analyzer

Table 4: Common Error Codes

If these or any other error codes appear on the display of the unit, contact Servomex for assistance. For an expedited turnaround, make sure that a System Data File is downloaded and attached if contact is made through email.

6.3.9.3 Fan Failure

The analyzer constantly monitors the condition of the cabinet exhaust fans. If hydrogen is entered as a gas in **GSF Setup**, and if the Hydrogen Safety System is installed, the analyzer will flash **H2 Safety Enabled** (alternating with **Isolated**) as shown in **Figure 52** if the system detects a failure in the exhaust fan circuitry. The analyzer will also isolate in this scenario and the user will be unable to restore any flow until the fan problem has been fixed.




Figure 52: Fan Failure Alarm

6.3.9.4 View GSF Log




Figure 53: View GSF Log Menu Selection




Figure 54: GSF Log

6.3.10

Analyzer Setup

The **Analyzer Setup** menu selection is shown in **Figure 22**. This menu selection access five submenus for analyzer setup.

6.3.10.1 Alarm Setup

The **Alarm Setup** menu selection is shown in **Figure 55**, below. This menu selection accesses submenus for **Moisture 1 -4**, **Temperature Range**, **Pressure Range**, and **System Error** alarms. The screens corresponding to these menu selections are shown in **Figure 56** through **Figure 59**.




Figure 55: Alarm Setup Menu Selection

The moisture analyzer includes a total of seven alarms. The four moisture concentration alarms dead bands are user-set.

Alarm Number	Function
1	Moisture Level 1
2	Moisture Level 2
3	Moisture Level 3
4	Moisture Level 4
T	Temperature
P	Pressure
S	System

Table 5: Alarm Codes

The temperature alarm indicates an out of specification ambient temperature condition in the analyzer cabinet. The pressure range alarm is related to the pressure in the gas path. Finally, system errors are monitored, and under certain conditions will trip alarms if enabled.

NOTE: When any hydrogen is included in the background gas matrix (see discussion regarding GSF on page 30 for additional information), an additional alarm is enabled to monitor the cabinet exhaust fan status. If the system detects a failure in the operation of the cabinet exhaust fans while operating in a hydrogen background, the system immediately isolates the moisture cell until the fan is repaired. A warning describing this condition will appear over the main display and the user will be unable to restore gas flow until the fan is repaired. See Fan Failure on page 54.

An alarm warning will overwrite the moisture level readout if an alarm condition exists. To acknowledge the alarm simply hit the **Enter** button and its number or letter will appear in the Alarm Status line above the display (refer to **Table 6: Alarm Codes**). This action will not clear the alarm. Only correction of the condition that caused the alarm will clear the alarm.

6.3.10.2 Alarm Setup




Figure 56: Alarm Setup Screen

The Setpoint value refers to the limit above or below which the alarm is triggered. The Trip command sets the above/below parameter. The dead band refers to the deviation from the nominal setpoint that the output value must exceed before an alarm is triggered. The Relay Assignment indicates to which relay the alarm is assigned.

Use the **Next** key to move from between fields and use the arrow keys (▲ and ▼) to change the highlighted selections and to enter numerical values. When done, use the **Next** key to move to the **Accept** button and hit the **Enter** key to return to the main display. Using the **ESC** key at any time will exit the screen making no changes and return to the main display.

6.3.10.3 Temperature Range Alarm Setup

Temperature Alarm Setup.vi

Moisture Sensor
Temperature Alarm Setup

Alarm
Off
On

Deadband
0.0

High Trip
21.0

Low Trip
5.0

Relay Assianed
N/U
1
2
3
4

Temperature inside Cabinet
30.0 °C

Press ESC to Cancel Accept

Figure 57: Temperature Alarm Setup

The system is constantly monitoring the ambient temperature in the analyzer cabinet. If enabled on the Temperature Alarm Setup screen, an alarm can be assigned to trip if the ambient temperature exceeds pre-set limits. The user may assign the temperature alarm to one of four relays.

Use the **Next** key to move from between fields and use the arrow keys (▲ and ▼) to change the highlighted selections and to enter numerical values. When done, use the **Next** key to move to the **Accept** button and hit the **Enter** key to return to the main display. Using the **ESC** key at any time will exit the screen making no changes and return to the main display.

6.3.10.4 Pressure Range Alarm Setup

The system is constantly monitoring the pressure in the gas path and the result is displayed on the front panel. If enabled on the Pressure Alarm Setup screen, an alarm can be assigned to trip if the pressure is outside pre-set limits. The user may assign the pressure alarm to one of four relays.




Figure 58: Pressure Range Alarm Setup

The limits are not user adjustable but are set automatically based on the background gases entered in the GSF screen. See page 30 for additional information on setting the background gases.

Use the **Next** key to move from between fields and use the arrow keys (▲ and ▼) to change the highlighted selections and to enter numerical values. When done, use the **Next** key to move to the **Accept** button and hit the **Enter** key to return to the main display. Using the **ESC** key at any time will exit the screen making no changes and return to the main display.

6.3.10.5 System Alarm Setup



Figure 59: System Alarm Setup


If an error is present for more than 30 minutes, a System Error Code will trip a System Alarm if configured to do so. See page 53 for additional information on System Error Codes.

NOTE: The 30-minute clock is delayed for 60 minutes on a “cold” system start up.

Use the **Next** key to move from between fields and use the arrow keys (▲ and ▼) to change the highlighted selections and to enter numerical values. When done, use the **Next** key to move to the **Accept** button and hit the **Enter** key to return to the main display. Using the **ESC** key at any time will exit the screen making no changes and return to the main display.

6.3.11

Analog Output Setup

**Figure 60: Analog Output Setup Menu**

The Analog Output Setup menu selection is shown in Figure 60. This selection accesses the Moisture Analog Setup screen shown in Figure 61.

The **Zero Point** corresponds to the lowest voltage or current output (0 VDC, 4 mA) that is sent to a recorder, while the Full Scale corresponds to the maximum voltage or current output (1/5/10 VDC or 20mA) that is sent. The **Full-Scale** set point (FS) is set from 0.002 ppm to 20.00 ppm.

Three ranges can be entered in this screen. The range of the primary Full Scale (FS) must be less than that of the Expanded Full Scale “A” (FS A) which must be less than that of the Expanded Full Scale “B” (FS B). The analyzer auto-ranges between the three outputs depending on the current analyzer reading.

Relay contacts can then be assigned to signal a change in range. If only one expanded range is required, rather than two, then the value of FS B should be set to equal the FS A value.

Moisture Analog Output Setup

Zero Point:	0.000 ppm	In	N/U
Full Scale:	0.100 ppm	Calibration	1
Expanded FS A:	1.000 ppm	Relay	2
Expanded FS B:	10.000 ppm		3
			4
Expanded	N/U	Output	Freeze
Range	1	During Cal	Update
A	2		
Relay	3	Expanded	N/U
	4	Range	1
		B	2
		Relay	3
			4

Press ESC to Cancel Accept

Figure 61: Analog Output Setup Screen

The **In Calibration Relay** can be enabled to signal that a zero calibration is in process. In addition, the user has the option to freeze the analog output or enable the analog output to update as the calibration progresses.

6.3.12 Graph Setup

The **Graph Setup Menu Selection** is shown in **Figure 62**, below. This menu selection accesses the **Graph Setup** screen in **Figure 63**.




Figure 62: Graph Setup Menu

The graph setup is used to adjust the time scale on the main data display of the analyzer. A specific time interval in minutes can be chosen for the X-axis on that display. The minimum acceptable time is 3 minutes. The information on the display represents current data and will show a history of moisture concentration based on the given time span to the present. The Y-axis of the main data display is auto-ranging.




Figure 63: Graph Setup Screen

6.3.13

Scaler Editor




Figure 64 Scaler Editor Menu

The **Gas Scaler Editor** screen, as shown in **Figure 66** enables the user to adjust the moisture reading by a factor of 0.3 to 3.0 for each gas included in the matrix.

Use the Next key to move from between fields and use the arrow keys (▲ and ▼) to change the highlighted selections and to enter numerical values. When done, use the Next key to move to the Accept button and hit the Enter key to return to the main display. Using the ESC key at any time will exit the screen making no changes and return to the main display.

Adjustments to the GSE screen will result in a warning as shown in **Figure 65** indicating the need to reset the **Gas Scale Factor** matrix. See **6.3.2 Set Up GSF** on page 30 for information on adjusting the **Gas Scale Factor** matrix.



Figure 65: GSF Warning




Figure 66 Scalar Editor Screen

6.3.14 Scheduled Reboot

The **Scheduled Reboot** menu selection is shown in **Figure 67**, below. This menus selection accesses the **Next Scheduled Reboot** screen shown in **Figure 68**.




Figure 67: Scheduled Reboot Menu Selection

Scheduled reboot allows scheduling of a periodic automatic reboot of the analyzer. The time interval is settable to between 1 and 5 weeks. Time is settable to any time using 24-hour clock notation. Navigation is accomplished using the **Next** and arrow buttons. The next scheduled reboot is displayed near the bottom of the screen as shown in **Figure 68**




Figure 68: Next Scheduled Reboot

6.3.15 Diagnostics Menu

The **Diagnostics** menu is used to select one of five diagnostic submenus. It is shown in **Figure 69** with the **Active Zero On/Off** selected. This selection access the **Active Zero On/Off** screen (**Figure 70**).




Figure 69: Diagnostics Menu Selection

6.3.15.1 Active Zero On/Off Screen



Figure 70: Active Zero Setup Screen

The **Active Zero Offset** feature is designed to automatically compensate for the analyzer's gradual zero baseline clean-up. This gradual clean-up is natural and occurs after a fresh startup or after a prolonged or abnormally high moisture exposure. This feature ensures that accurate low ppb H₂O readings can be made as soon as possible after initial start-up, or after a high H₂O upset event. It is like the User Zero Offset feature in that a small positive offset is added to the analyzer H₂O readings (display and output) to compensate for the long-term downward trending in the readings during clean-up. The **Active Zero Offset** provides an automatic addition of offset that occurs in miniscule steps, and within set guidelines, corresponding to predictable behavior during cleanup.

When **Active Zero Offset** is enabled through the Diagnostics Menu, the **User Zero Offset** feature is disabled and vice versa. While the **User Zero Offset** feature requires the user to enter a fixed positive offset value to accommodate the baseline clean-up, the **Active Zero Offset** does so automatically, and only when necessary.

The current **Active Zero Offset** value is shown in the **Check/Adjust Zero** screen as shown in **Figure 30**. It starts at a value of 0.00 ppb when the analyzer is first turned on, and then increments automatically as the analyzer applies offset to the readings. After each User Zero Calibration, the **Active Zero Offset** value is reset to 0.00 ppb and then automatically increments again as needed.

The **Active Zero Offset** is designed to operate when the zero baseline is falling at a rate less than 0.1 ppb/hr. as would be the case after 1-2 weeks of initial operation. If user calibrations are performed sooner, the H₂O readings may be decreasing too rapidly for the Active Zero Offset feature to operate properly and inaccurate H₂O readings may result.

If the **Active Zero Offset** value reaches 5 ppb a warning message CAL ZERO will flash in the system status block on the display instructing the user that a zero calibration should be performed. The maximum amount of offset that can be applied by this feature is 5 ppb. Any further downward trend (baseline clean-up) exceeding -0.3 ppb will result in inaccurate readings until the next user calibration is performed resetting the **Active Zero Offset** value to zero.

If the **Active Zero Offset** feature is turned off, the **User Zero Offset** value will appear in its' place in the Check/Adjust Zero menu. The previous **User Zero Offset** value (if any) will reappear and immediately be applied to the live display readings. Likewise, if the **Active Zero Offset** feature is on, then its value (if any) will appear and immediately be applied to the live readings.

6.3.15.2 Test Relays

The **Test Relays** menu selection is shown in **Figure 71**, below.




Figure 71: Test Relays Menu Selection

The **Test Relays** screen, as shown in **Figure 72**, is used to assure that the relay outputs are functioning properly. When the Test Relays screen is selected, the **NEXT** key is used to move to the number field where the arrow keys (▲ and ▼) are used to choose the appropriate relay number. The **NEXT** key is then used to move to the **Activate/Deactivate** field where the arrow keys (▲ and ▼) are used to toggle between the two options. The **NEXT** key is then used to move to the **Apply** field where the **Enter** key is hit to change the relay state. This process can be repeated as often as needed. When done, the **NEXT** key is used to move to the **Done** field and the **Enter** key is hit to leave the screen. The condition of the relays before the test is restored when the test is concluded. See the section on relay ports found on page 26 for additional information.




Figure 72: Test Alarm Relays Screen

6.3.15.3 Test Analog Voltage Output

The **Test Analog Outputs** menu selection is shown in **Figure 73**, below. This menu selection accesses the **Test Analog Outputs** screen shown in **Figure 74**. This screen is used to calibrate the analog recorder output.




Figure 73: Test Analog Outputs Menu Selection

When the Test Output screen is displayed, the **NEXT** key is used to move to the percentage field where the arrow keys (▲ and ▼) are used to choose the appropriate setting. The **NEXT** key is then used to move to the **Apply** field where the **Enter** key is hit to set the analog output to the selected value. The analog output response should match the value that was entered. For example, if 80 percent is entered for the percent full scale level, and the analog output is set for 0 to 10 VDC, the analog output is 8.000 VDC. This process can be repeated as often as needed. When done, the **NEXT** key is used to move to the **Done** field and the **Enter** key is hit to leave the screen. The condition of the analog output before the test is restored when the test is concluded. See the section on analog outputs found on page 25 for additional information.




Figure 74: Test Analog Voltage Output Screen

6.3.15.4 Signal Monitor

The **Signal Monitor** menu selection is shown in **Figure 75**, below.




Figure 75: Signal Monitor Menu Selection

The **Signal Monitor** screen (**Figure 76**) depicts 18 system parameters, which are defined in **Table 6**. Values for parameters are unique to each system. In the event of a system error, the parameters can be used for diagnostic purposes. Contact Servomex for assistance in using signal monitor parameter data as a diagnostic tool for maintaining system performance.




Figure 76: Signal Monitor Screen

Parameter	Definition
P1	Peak Acquisition
P2	Sample Detector DC
P3	P Chamber
P4	Cell Temp (= T gas)
P5	T Laser
P6	T Laser Set Point
P7	Baseline Adjust In
P8	Fan Condition
P9	Ambient Temp
P10	Fit RMS Error (filtered)
P11	Original SDC
P12	Original P Chamber
P13	Original Cell Temp
P14	Original T Laser Set Point
P15	Original Fit RMS Error (filtered)
P16	Raw PPB SD (Past 24 hours)
P17	Max. Raw PPB (Past 24 hours)
P18	Min. Raw PPB (Past 24 hours)

Table 6 Signal Monitor Parameter Definitions

6.3.15.5 Cell Pressure Calibration

The **Cell Pressure Calibration** menu selection is shown in **Figure 77** below. This menu selection accesses the **Cell Pressure Calibration** screen show in **Figure 78**.




Figure 77: Cell Pressure Calibration Menu Selection

Cell pressure calibration allows calibration of the cell pressure transducer. Although the transducer is calibrated at the factory, if the analyzer is located at a relatively high altitude or has had the transducer replaced, a calibration check should be performed. The default target is 760 Torr (nominal seal level pressure), however, as atmospheric pressure varies with elevation and weather conditions, it is recommended that the nearest weather station be used to determine the current ambient pressure. The vacuum source (pump or aspirator) should be disconnected, and gas flow stopped when performing calibration. To navigate, press the “Next” button to highlight the gain adjustment box. Move the cursor with the left and right arrows as needed. Change the selected digits using the up and down arrows. This screen also allows a manual offset to be applied to the displayed pressure. This feature is generally for testing purposes and should not be used during routine calibrations.




Figure 78: Cell Pressure Calibration Screen

6.3.15.6 Manual Cell Pressure Override

The Manual **Cell Pressure Override** menu selection is shown in **Figure 79**: Manual Cell Pressure Override Menu Selection below. This menu selection accesses the **Manual Cell Pressure Override** screen show in **Figure 80**.




Figure 79: Manual Cell Pressure Override Menu Selection

Manual cell pressure override allows the use of either the pressure reading provided by the transducer or a manually entered pressure to determine moisture concentration. Navigation is performed using the **Next** and arrow buttons. The default setting is **Sensed**, meaning the analyzer is using the value provided by the pressure transducer. Selecting **Entered** enables the user to manually enter the pressure to be used by the analyzer. Care should be taken when doing this. A large error in moisture readings will occur if the entered reading is not close to the actual pressure.




Figure 80: Manual Cell Pressure Override

6.3.16 Power Up Default (applicable to analyzers with iso valves)

The **Power Up Default** menu selection is shown in **Figure 81**, below. This menu selection accesses the **Power Up Default** screen shown in **Figure 82**.




Figure 81: Power Up Default Menu Selection

The Power Up Default selection allows the user to set the default state of the isolation valves after analyzer power up. This feature is available only on analyzers that have the isolation valves.

Use the **NEXT** key to move from field to field. Use the up and down arrows (▲ and ▼) to move between the selections within the field. When done use the **NEXT** key to move to the **Accept** field and hit the **Enter** key. Pressing the **ESC** at any time will make no changes and will return to the main data display.

NOTE: If the analyzer auto-reboots due to a system error, the power-up defaults are ignored, and the analyzer returns to the mode of operation found immediately before the error. See page 53 for additional information on system error codes.




Figure 82: Power Up Default Screen

6.3.17

Date/Time

The **Date/Time** menu selection is shown in **Figure 83**, below. This menu selection accesses the **Date/Time Setup** screen shown in **Figure 84**.




Figure 83: Date/Time Menu

The Date/Time Screen is used to set various calendar and clock related parameters.

The **Next** key is used to move from field to field, and the arrow keys (▲ and ▼) are used to change the numerical digits and units.

When in the Time Zone field, the left and right arrow keys toggle through the various options. When done, the use the **Next** key to move to the **Accept** field and press the **Enter** key to confirm.

Select **Change date and time** and press the **Enter** key to access the **Date/Time Edit** screen (Figure 85). Use the arrow keys to change the date and time in this screen. When done, select **OK** and then press **Enter** key to confirm.

Pressing the **Esc** key at any time will exit the screen with no changes and return the user to the main data display.

NOTE: The time is not automatically adjusted for daylight savings and must be changed manually.




Figure 84: Date/Time Setup Screen




Figure 85 Date/Time Edit Screen

6.3.18

Communications

The **Communications** menu selection is shown in **Figure 86**, below. This menu selection access the **Communications Setup** screen shown in **Figure 87**.




Figure 86: Communications Menu Selection

The Communications screen is used to set parameters related to serial PC communications. The **Next** key is used to move from field to field, and the arrow keys (▲ and ▼) are used to change the numerical digits as well as to select the baud rate. When done, the **Next** key is used to move to the **Accept** field and the **Enter** key is hit. See page 24 for additional information.




Figure 87: Communications Setup Screen

6.3.19

Download System Data

The **Download System Data** menu selection is shown in **Figure 88**, below.




Figure 88: Download System Data Menu

The contents of the internal system data files can be downloaded to a USB memory stick, and the files can either be mailed or e-mailed to Servomex for evaluation. Install a memory stick into the external USB socket located behind the front door and on the left side of the analyzer. After selecting **Enter** from the **Download System Data** menu, **Figure 89** is displayed.

The system downloads data for ten days ending with the date set on the screen as shown in **Figure 89**. The current date is set automatically but can be changed to capture activity at a specific time other than the last 10 days. To change the date, use the **Next** key to move between fields and use the arrow keys (▲ and ▼).

Press **Enter** and the download process will begin, and a time bar will appear as in **Figure 92**. The process should only take a minute or two and when complete control will be returned to the user.




Figure 89: System Data Download



Figure 90: Media Warning

If there is insufficient space available on the media a warning will appear as in **Figure 90**. If Drive D is not ready, a message will appear as in **Figure 91** below.



If not ready message appears, select OK, check that USB is inserted correctly and try again.




Figure 91: Not ready Drive d:




Figure 92: Download Timeline

The file name is automatically created and includes the date and time that the data was recorded as well as the serial number of the analyzer. All files are then automatically compressed and loaded as one file on the memory stick, which can then be used to forward the information to Servomex for evaluation.

6.3.20 System Info

The **System Info** menu selection is shown in **Figure 93**, below. This menu selection accesses the **System Info** screen shown in **Figure 94**.




Figure 93: System Info Menu

The **System Info** screen (**Figure 94**) gives the user information regarding the configuration of the analyzer as well as the version of firmware currently installed. The **Service Menu** is password protected. Do not use this function unless instructed by Servomex.




Figure 94: System Info Screen

6.3.20.1 Firmware Upgrade

While the **Firmware Upgrade** box is highlighted, pressing the **Next** key will bring up the Firmware Upgrade dialog box as in **Figure 95** below.

Place the memory stick with appropriate files in the USB port located behind the front door. When ready, press the **Yes, Proceed** key and follow the instructions. At the conclusion of the upgrade the analyzer will automatically reboot.




Figure 95: Firmware Upgrade

7. Sample Gas Preparation and Delivery

7.1 Introduction

It is important to monitor sample pressure to ensure it does not vary from the specified range for the background gas being supplied to the moisture analyzer. Pressure and background gas composition significantly affect system performance. Flow rate also affects analyzer performance, but is less critical.

7.2 Sample Pressure

The DF-75SGMAX analyzer operates with sample gases at ambient pressure (nominal 760 Torr.)

7.3 Flow Rate Effects on Sensor Performance

Assuming a leak-free system, flow rate changes will have minimal effects on the performance of the moisture cell.

7.4 Background Gas Effects on Indicated Flow Rate

If the molecular weight of the background gas is much different from N₂, the flowmeter reading is not accurate. The bypass loop flowmeter mounted on the rear of the DF-700 series Moisture Analyzer is calibrated for use in air (or N₂). Most other gases have molecular weights within ± 25 percent of air. Since the required flow rate is not extremely critical, most gases produce reasonably correct readings. The exceptions are light gases such as Helium and Hydrogen, whose flow rates should be set to approximately one-third that of Nitrogen.

7.5 Flammable Sample Gas

The only flammable gas that the DF-745SGMAX can be used with is hydrogen. If used with hydrogen, the optional Hydrogen Safety System should be installed, or the analyzer must be installed in a purged environment.

There is nothing within the analyzer sample system that can ignite a flammable sample gas. However, it is critical to ensure that the sample gas does not escape from the sample system into the analyzer enclosure, or the room, where ignition is possible.

The analyzer enclosure can be purged with nitrogen, or the entire Analyzer can be mounted in a purged enclosure, so that any sample gas that escapes the plumbing is diluted.

8. Service

8.1 Return Material Authorization number

If the analyzer is being returned to the factory, call the Servomex Business office to obtain a **Return Material Authorization number**. Clearly mark the Return Material Authorization number on the outside of the shipping container and on the packing list.

The analyzer should be returned (freight prepaid) to:

Servomex Corporation
4 Constitution Way
Woburn, MA 01801-1087

8.2 Cautions Related to Maintenance

CAUTION: HAZARDS AFFECTING SERVICE PERSONNEL



Service personnel perform activities that expose them to hazards, such as disconnecting pressurized sample and vent lines, replacing power cords, and replacing fuses. Read the **Safety Summary** at the beginning of this manual before performing any maintenance on this analyzer. Also, read the cautions and manual instructions that are included in the section that covers the specific maintenance operation.

CAUTION: RETURNING THE ANALYZER TO SERVICE



Following maintenance, ensure that the analyzer is in a safe state. Close the AC inlet fuse door. Fasten the top or left side analyzer covers, using all screws. Ensure that inlet and outlet valves are in the appropriate position for normal operation and the power cord is properly plugged into a grounded outlet. Ensure that the AC power cord and plumbing are routed to avoid a tripping hazard.

8.3 Maintenance

8.3.1 Storage

Refit any protective plastic covers and place the analyzer and any associated equipment in its original packaging before storage. Alternatively, seal it inside a waterproof plastic bag, sack, or storage box. Store the analyzer and any associated equipment in a clean, dry area. Do not subject it to excessively hot, cold, or humid conditions.

If the analyzer is to be stored for extended periods of time, be sure that the temperature of storage location does not exceed 50° C (122° F). Storage in direct sunlight can cause temperatures to exceed the recommended limits even though ambient temperatures may be below the maximum temperature.

8.3.2 Fuse Replacement

CAUTION: HAZARDOUS AC



Before replacing fuses, turn off the analyzer power switch and allow the analyzer to shut down normally. Disconnect the analyzer power cord from the AC inlet.

Main fuses are in the AC inlet module. Facing the AC inlet there is a tab, on the left side of the inlet, which is pried to open the fuse cover. Gently pry the tab with a straight blade screwdriver, being careful not to damage the tab. The cover will flip open to expose a red fuse carrier. Use the screwdriver to pry the fuse carrier out of the AC inlet. Both sides of the AC line are fused. Note the position of the two fuses. This fuse carrier can accommodate different length fuses, and it is important to install the new fuses in the same position as the original fuses.

Replacement fuses may be purchased from Servomex or purchase the following:

- For 110 VAC operation -- 250 VAC, 6.3A, 5X20 mm, IEC 60127-2 Sheet 3, such as Littelfuse 021806.3HXP.
- For 220 VAC operation -- 250 VAC, 3.15, 5X20 mm, IEC 60127-2 Sheet 3, such as Littelfuse 02183.15HXP.

8.3.3 Moisture Cell Maintenance

The moisture cell does not require any periodic maintenance or inspection. If you observe that the moisture cell is not performing correctly, or if moisture-cell related error codes are displayed that are not clearable, contact Servomex.

8.3.4 Vacuum Pump Maintenance

If equipped, the vacuum pump requires periodic maintenance every 6 – 12 months to maintain proper pressure. The interval will vary depending on characteristics of the installation. A good working pump will pull down to approximately 115 torr or less when the moisture cell is isolated. When pump performance degrades to the point that proper pressure cannot be attained, a significant improvement in performance can often be achieved by cleaning of the cylinder and piston assembly. A rebuild kit is available from Servomex to return the pump to original specifications.

8.3.5 Power Cord Replacement

CAUTION: POWER CORD SAFETY



The power cord provides the protective earth connection. Do not use a power cord without a protective earth wire. Do not use a power cord with inadequate ratings (less than 120 VAC, 10 Amps for 110 VAC, or less than 240 VAC, 10 Amps for 220 VAC).

CAUTION: HAZARDOUS AC

Before replacing the power cord, turn off the analyzer power switch and allow the analyzer to shut down normally. Disconnect the analyzer power cord from the AC inlet.

For 110 VAC operation, the power cord is a Volex 17251 10 B1. Any equivalent three-wire power cord may be used provided it has a protective ground wire and conforms to the following specifications: 120 VAC, 10 A, 18 AWG, SVT jacket, with a NEMA 5-15P plug and an IEC 60320 C-13 receptacle for connection to the analyzer. Make sure that the cord has the necessary agency approvals required for your country.

For 220 VAC operation, any three-wire power cord may be used provided it has a protective ground wire and conforms to the following specifications: 250 VAC, 10 A, with a plug appropriate for your country and an IEC 60320 C-13 receptacle for connection to the analyzer. Make sure that the cord has the necessary agency approvals required for your country.

8.4 Troubleshooting Guide for the NanoTrace Analyzer

8.4.1 Moisture Troubleshooting

The DF-700 series moisture analyzer constantly monitors analyzer performance. In the event of a failure, a system alarm will be displayed on the front panel. In addition, the failure will be logged in the System Error Log (see **6.3.9.2 System Error Codes**). In the event of a system alarm, download system history as described in **6.3.7 Data History Routine** and provide the downloaded file to Servomex for assistance in troubleshooting.

8.5 Shipping

If it is necessary to return the analyzer to the factory or ship it to another location, follow the packaging and shipping procedure below to prevent damage to the analyzer during shipment.

1. If analyzer has isolation valves, isolate the analyzer moisture cell by following the steps in **4.2.2 Isolate Analyzer**.
2. Turn off the power switch. Disconnect any source of AC power from the analyzer.
3. Disconnect all external electrical connections (alarms, data output, communications etc.) and mark for reattachment later.
4. Mark each for reattachment later.
5. Ensure that all internal components are adequately secured and put the analyzer in its original container.

If the analyzer is to be returned to the factory, call the nearest Servomex Business office to obtain a **Return Material Authorization number**. Clearly mark the Return Material Authorization number on the outside of the shipping container and on the packing list.

The analyzer should be returned (freight prepaid) to:

Servomex Corporation
4 Constitution Way
Woburn, MA 01801-1087

8.6 Disposal

Dispose of the analyzer and any associated equipment safely, and in accordance with all of your local and national safety and environmental requirements.

*Hint: If you send the analyzer to Servomex or your local Servomex agent for disposal, it must be accompanied by a correctly completed decontamination certificate and a Return Authorization Number (RAN) Refer to **8.5 Shipping**).*

8.6.1 Disposal in accordance with Waste Electrical and Electronic Equipment (WEEE)

The label shown in **Figure 96** The WEEE label) is fitted to the analyzer.



Figure 96 The WEEE label

This label identifies that:

- The analyzer is considered to be within the scope of the Waste Electrical and Electronic Equipment (WEEE) Directive.
- The analyzer is not intended for disposal in a municipal waste stream (such as landfill sites, domestic recycling centers and so on), but must be submitted for material recovery and recycling in accordance with the local regulations which implement the WEEE Directive.

For additional information and advice on the disposal of the analyzer in accordance with the requirements of the WEEE Directive, contact Servomex or your local Servomex agent.

Servomex Group Limited, Jarvis
Brook, Crowborough, East Sussex,
TN6 3FB, England
Tel: + 44 (0)1892 652181
Fax: +44 (0)1892 662253
Email: Europe_sales@servomex.com

9. Theory of Operation

9.1 The Moisture Measurement




Figure 97 Schematic of Moisture Cell

9.1.1 Moisture and the IR Spectrum

The NanoTrace Moisture Analyzer uses infrared (IR) absorption as its detection method. IR absorption is a subset of the overall field of spectroscopy, which concerns the interaction of light and matter. A conceptual drawing of the type of moisture cell used in the analyzer (called a Herriott cell for physicist Donald R Herriott) is shown in **Figure 97**. The cell includes a laser light source, an enclosed cavity through which the sample gas passes, a system of mirrors, and a light sensor.

Absorption spectroscopy is based on the absorption of energy that occurs when electromagnetic waves (specifically light, in this case) collide with matter. In the infrared spectral region, the wavelengths of light overlap with absorptions created by molecular vibrations. IR absorption is often used in evaluating molecular composition based on “fingerprints” of these absorptions over a wide wavelength range, but if the strength of a single vibrational absorbance is known, a single wavelength can be used to determine the amount of a particular substance. This is the method used by Servomex 700-Series analyzers. It should be noted that there are factors other than changing moisture concentration that affect the amount of light that reaches the sensor. The moisture cell mitigates this problem by scanning across a narrow wavelength range and continuously comparing the intensity of light at the interfering frequency with a non-interfering frequency.

9.2 Absorption Spectroscopy

The relationship that defines absorption spectroscopy is known as Beer's Law. Beer's Law equates, in rigorous terms, the concentration of any absorbing molecule based on absorbed light intensity at a particular wavelength, given knowledge of the molecule's absorption strength and the "path length" of the sample medium.

Many are familiar with Beer's Law as it is conventionally used in analytical laboratories:

$$A_{\lambda} = \log(1/T)_{\lambda} = \log(I_0 / I)_{\lambda} = 2.303 \times \varepsilon_{\lambda} \times b \times C$$

Definitions:

$A_{\lambda} \equiv$ Absorbance at wavelength λ

$T \equiv$ Transmittance

$I_0 \equiv$ Reference Intensity of Light

$I \equiv$ Measured Intensity of Light after Absorption

$\varepsilon_{\lambda} \equiv$ Molar Absorptivity at wavelength λ

$b \equiv$ Path length

$C \equiv$ Molar concentration

In this embodiment, a solution with a broad absorbance band is dialed to a wavelength within the band, where a substance's molar absorptivity is known, and the concentration of that substance is determined. The substance is usually a liquid solution, placed in a 1 cm cuvette, and the concentration is expressed in moles/liter.

This same Law can have units reassigned to determine absolute numbers of molecules per cubic centimeter, useful in gaseous measurements:

$$A_{\lambda} = \ln(I_0 / I)_{\lambda} = \sigma_{\lambda} \times b \times N$$

$\sigma_{\lambda} \equiv$ Molecular Cross Section ($\text{cm}^2/\text{molecule}$) at wavelength λ

$N \equiv$ Molecular Density ($\text{molecules}/\text{cm}^3$)

The values σ and ε are related primarily by Avogadro's Number. If the molecular density of an absorbing substance in gas, such as moisture, is known, it can be compared to the number of molecules in an ideal gas, resulting in a report of parts per billion (ppb). Since the measurement is calculated per unit volume instead of per molecule count, this concentration is more specifically called parts per billion in volume (ppbv).

10. Specifications

10.1 Moisture

Lowest Detection Level (LDL): 5 ppb @ Constant Conditions

Accuracy: Greater of $\pm 5\%$ of reading or ± 5 ppb @ constant conditions

Speed of Response: Less than 3 minutes to reach 90 percent of an upward step change

Upset Recovery Time: Less than 5 minutes from a high ppb upset to within 50 ppb of the previously stable reading.

Range: 0-100 ppm (smallest recommended range: 0-1ppm)

Background Gas Compatibility N₂, O₂, H₂, He, Ar, CO₂, CO, SF₆, Ne, Kr, Xe, N₂O, CF₄, C₂F₆, C₃F₈, C₄F₈, NF₃, and blends of up to eight of supported gases.

Cell Pressure: Ambient pressure

10.2 General

Construction: NEMA 1 (IP10), 19-inch rack mount

Dimensions: 19 inch (48.3cm) wide x 10.5 inch (26.7 cm) high x 22.5 inch (57.2 cm) deep

Display: 10.1-inch LCD Color 1024X600)

Weight: 73 pounds (33.11 kg)

Required Sample Conditions: oil free, non-corrosive, non-condensing, filtered for particulates to 2 μ m

10.3 Environmental:

Altitude: 2000 meters (6500 feet), max

Ambient Operating Temperature: 10° C to 40° C (50° F to 105° F)

Humidity: 5% relative humidity non-condensing

General: Indoor use only in a clean, dry, non-hazardous area of pollution degree 2, as defined in EN 60664-1

Storage Temperature: Not to exceed 50° C (122° F)

10.4 Gas Routing

Gas Inlet, Sample:

Flow Rate: 1.3 slpm total (moisture cell: 1 slpm, bypass outlet: 0.3 slpm)

Inlet Pressure: 18 to 150 psig (1.24 to 10 barg, 124.1 to 1034.2 kpag)

Vent: Vent to atmosphere. Vent pressure limits: -2 to 2 psig, (-0.14 to 0.14 barg, -13.7 to 0.13 kpag)

Gas Inlet, Pneumatic:

Gas: N2 or air

Inlet Pressure: 70 to 100 psig (4.8 to 6.9 barg, 482.6 to 689.4 kpag)

Span Gas: Not required.

Zero Gas: Optional. Recommended if operating near LDL

Gas Path Construction Materials: 316 Stainless Steel, 304 Stainless Steel, Indium, Passivated Steel, Hastelloy C22, Gold, Invar, Acrylic, Buna-N, Sapphire, PCTFE, Polypropylene,

10.5 Output Signals & Status Indicators

Alarm Relays: Four non-latching, independently assignable to moisture alarms. SPDT contacts rated at 1 Amp @ 30 VDC resistive load. Fail safe action upon loss of power to alarm condition. Not designed to switch AC power.

Analog Output: 0-1, 0-5, or 0-10 VDC (minimum load resistance: 1K), User Scalable 0-2 ppb to 0-20 ppm Moisture.

Visual Alarm Status Indicators: Four Moisture levels, Temperature, System Error, Zero Verification or Calibration in Process, Fan Failure.

Dual Scale Range: Two user selectable secondary analog output ranges for rescaling the output to display screen once the primary range is exceeded

Isolated 4-20 mA DC: 1K ohm loop resistance max, (28V Compliance voltage provided).

Digital Communications: Two-way RS-232 or RS-485 set at the time of order.

10.6 Laser

Max Power: 50mW/30mW

Beam Divergence: 436 mrad

Pulse Pattern: CW

Analyzer Product Classification: Class 1 (per BS EN 60825-1:2014 and 21CFR 1040.10 and 1040.11 except for deviations pursuant to Laser Notice No. 56, 5 May 2019)

10.8 Power

100 to 120 VAC, 50/60 Hz, standard, 5 Amps

200 to 240 VAC, 50/60 Hz (optional), 2.5 Amps

Warranty: One year from ship date on the entire instrument. See **Chapter 10 Warranty**.

10.9 Compliance

EMC: (EMC) Directive 2014/30/EU

Electrical Safety: EU Low Voltage Directive; IEC 610110-1 Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use

11. Warranty

Servomex warrants each instrument it manufactures to be free from defects in material and workmanship at the F.O.B. point specified in the order, its liability under this warranty being limited to repairing or replacing, at the Seller's option, items which are returned to it prepaid within one (1) year from delivery to the carrier and found, to the Seller's satisfaction, to have been so defective.

In no event shall the Seller be liable for consequential damages. NO PRODUCT IS WARRANTED AS BEING FIT FOR A PARTICULAR PURPOSE AND THERE IS NO WARRANTY OF MERCHANTABILITY. Additionally, this warranty applies only if: (i) the items are used solely under the operating conditions and in the manner recommended in the Seller's instruction manual, specifications, or other literature; (ii) the items have not been misused or abused in any manner or repairs attempted thereon; (iii) written notice of the failure within the warranty period is forwarded to the Seller and the directions received for properly identifying items returned under warranty are followed; and (iv) with return, notice authorizes the Seller to examine and disassemble returned products to the extent the Seller deems necessary to ascertain the cause of failure. The warranties stated herein are exclusive. THERE ARE NO OTHER WARRANTIES, EITHER EXPRESSED OR IMPLIED, BEYOND THOSE SET FORTH HEREIN, and the Seller does not assume any other obligation or liability in connection with the sale or use of said products.

Index

4

4-20mA Outputs, 25

A

about this manual, ix

scope of manual, ix

symbols, ix

Active Zero On/Off, 66

Add Location, 49

Alarm Setup, Moisture, 57

Analog Signal Outputs, 25

Analyzer

Specifications

General, 93

Moisture, 93

Warranty, 96

Analyzer Shut Down, 20

B

Back Flow Prevention System, 18, 38

C

case purge, 99

Check/Adjust Zero, 41

Communications, 80, 81

D

Data Download, 47

Data History Routine, 44

Delete Location, 50

Display Setup, Moisture, 62

disposal, 90

Download System Data, 81

E

Error Codes, 53

External Devices, Connecting to

4-20mA Outputs, 25

Analog Signal Outputs, 25

Relay Ports, 26

Serial Communications, 24

External Devices, Connecting to, 24

F

Fan Failure, 54

Firmware Upgrade, 85

G

Gas Flow and Pressure Regulation, 17

H

Hydrogen service, 99

I

important information, ii

In Calibration Relay, 61

Installation and Setup, 11

Isolate Analyzer, 30, 37, 40, 41, 44, 47, 50, 51, 53, 55, 56, 61, 64, 65, 67, 69, 71, 73, 74, 76, 78, 80, 81, 83

K

Key Lock, 23

M

Manual Zero, 43

Moisture Alarm Setup, 57

Moisture Calibrate Menu, 40

Moisture Data History Routine, 45

Moisture Display Setup, 62

Moisture Temperature Range Alarm, 58

N

nitrogen case purge, 99

P

Power Up Default, 77

Pressure Range Alarm Setup, 58

Pump case purge, 99

R

RAN, 90

Relay Ports, 26

Restore Sample Gas Flow, 40

Return Authorization Number, 90

S

Sample Flow Rate, 86

Sample Gas Flammability, 86

Sample Gas Preparation and Delivery, 86

Background Gas Effects, 86

Flow Rate Effects

Sensor Performance, 86

Index

Sample Gas Flammability, 86
Service, 87
 Maintenance
 Storage Conditions, 87
 Return Material Authorization number, 87
 Shipping, 89
 Return Material Authorization
 number, 87, 89
Set Location, 48
Shipping, 89
Signal Monitor, 71
symbols, ix
System Alarm Setup, 59
System Error Codes, 53

T

table of figures, vi
Test Analog Voltage Output, 70
Test Relays, 68
Theory of Moisture Measurement, 91

U

User Interface, 27
 Data Display Screen, 27
 Keypad, 27
 Menu Structure, 28
User Zero Offset, 43

V

View Location, 49
View Moisture Logs, 50

W

Warranty, 96
WEEE Directive, 90

Z

Zero Log, 51

Appendix A – Hydrogen Safety System

The **Hydrogen Safety System** is designed to safeguard the DF-745SGMAX from explosion hazards when operating on hydrogen sample gas under normal pressure and flow conditions as described in this manual. The instrument chassis and the remote pump, if equipped, are both protected by maintaining a safe condition within their respective enclosures.

If equipped with this option at the time of shipment from the factory, the analyzer will be set for Hydrogen operation through the GSF screen, and the Hydrogen Safety Service System will automatically be enabled.

Instrument

- **Air circulation fans** pull air in through the front door (and vent through the rear) of the chassis. The fans are rated at 50 cfm (130 cfm max) which will maintain the internal chassis space below the lower explosive limit (LEL)¹.
- **Sample delivery interlock valve** blocks flow of sample gas from entering the instrument chassis under various conditions as described below. It consists of a normally closed pneumatically actuated UHP springless diaphragm valve positioned on the sample inlet bulkhead.
- **Analyzer case purge valve** introduces nitrogen purge flow into the chassis upon closure of the sample delivery interlock valve. It consists of a normally open pneumatically actuated valve which feeds a customer supplied and regulated source of nitrogen purge gas through a bulkhead on the rear of the cabinet.²
- **Instrument controls and logic**³ manage and actuate the above-mentioned components to maintain safe operating conditions.

Vacuum Pump

- If equipped with a vacuum pump, an enclosure equipped with a nitrogen case purge is provided to maintain a safe condition. In addition, an interlock is provided to detect and react to loss of Nitrogen purge flow. Refer to **Figure 98**.

Installation

If equipped with this safety system, the analyzer installation procedure is modified as follows:

- If the analyzer has the optional vacuum pump, the pump enclosure must be mounted to a nearby wall or inside the rack shared by the analyzer.
- The sample gas inlet connection is made to an interlock valve mounted at the sample gas inlet on the rear of the analyzer. NOTE: The installation of this option changes the sample inlet connection to female VCR, from male.
- The sample gas outlet is connected by a ¼ inch metal tube from the analyzer sample outlet to the aspirator. If the analyzer has the optional vacuum pump, the sample gas outlet is connected by a ¼ inch metal tube from the analyzer sample outlet to the needle control valve (supplied loose), and then to the sample inlet on the pump enclosure.
- A customer supplied and regulated source of nitrogen purge gas is connected to both the ¼ inch compression fitting on the case purge valve mounted on the rear of the analyzer, and, if equipped with the vacuum pump, to the 1/8-inch compression purge inlet fitting on the vacuum

pump enclosure. Using the customer supplied regulator, adjust the flow into both enclosures to approx. 30 scfh as indicated on a rotameter.

- If required, the pump enclosure sample outlet and/or analyzer sample outlets need to be connected by ¼ inch metal tube and compression fittings to an appropriate exhaust system.
- The pump power cord is connected from the rear of the analyzer to the rear of the pump enclosure.

Operation

- Before power is applied to the analyzer, the moisture cell sensor is isolated as the internal inlet and outlet valves are closed. In addition, the external sample interlock valve is closed, and the case purge valve is opened, allowing customer supplied and regulated Nitrogen gas to purge the analyzer cabinet.

When power is applied, and only after the system has verified proper exhaust fan operation, the system automatically opens the external sample interlock valve to enable gas flow and the case purge valve is closed.

NOTE: The backflow prevention screen (**Figure 25**) will be displayed and the user must hit **ESC** to acknowledge the screen and remove it.

After approx. 5 minutes, the Warming Up indication will disappear, and the user is then free to open the internal inlet and outlet valves to the moisture cell.

- On analyzer shutdown, or in the event of a power failure or fan failure, the moisture cell is automatically isolated with internal valves, the external sample interlock valve closes to block sample flow from entering, and the purge valve opens to allow Nitrogen purge into the analyzer enclosure.
- The pump case purge system continuously feeds a customer-regulated supply of nitrogen into the pump enclosure to: (1) maintain the hydrogen level well below the maximum safe level (4% for hydrogen) if the pump diaphragm fails and leaks sample gas into the enclosure; and (2) maintain appropriate flow for adequate pump cooling. The purge flow rate is set to 30 scfh as indicated on a rotameter mounted on the side of the pump enclosure. An in-line flow switch will trip at a flow rate of less than 26 scfh assuring adequate flow. Loss of purge flow breaks the contacts in the flow switch, which in turn trips a mercury relay that removes electrical power to the pump.

CAUTION



The purge flow out of the pump enclosure may contain sample gas and should be appropriately vented.

CAUTION

After passing through the moisture cell and orifice, the sample is under vacuum and any leak in the system would result in ingress of ambient rather than release of sample gas to ambient. Accordingly, the sample outlet should be vented appropriately to assure adequate dilution.

CAUTION

The Hydrogen Safety System is designed to be safe as provided by the factory when operating as described in the Operating Instruction Manual. DO NOT make additional penetrations in the enclosure. If there is a need to do so, please contact the factory. Additional penetrations made in the enclosure (or failure to properly fasten the enclosure door) will allow additional influx of ambient oxygen and may defeat the case purge safety mechanism. A pressure relief valve is installed to prevent over-pressurization of the enclosure.

CAUTION

Do not open the pump enclosure door unless AC power is shut off.

CAUTION

The operator is obligated to assure proper operation of the analyzer air flow system as designed. Do not impede air flow at the inlet in the front door or at the exhaust fan outlets on either side of the cabinet in the rear.

NOTES

¹ For hydrogen, which has a lower explosive limit (LEL) of 4%, the maximum allowable influx in event of an internal leak would be 120 scfh, whereas the normal flow as defined in the Operating Instruction Manual is about 1/10th that at approximately 14 scfh.

² Use of the analyzer case purge is at the customer's discretion. It most likely is not necessary since in the event of sample delivery interlock, the sample feed is blocked from entering the chassis and the only open flow path through the system is dissipation of remaining sample in the bypass, which would be vented externally, or in the event of an internal leak, would be volumetrically insignificant.

³ The pneumatic control for the sample interlock and case purge valves is provided by an internal 12 VDC solenoid.

NOTE: On the DF-745SGMAX only: Relay 8 has been permanently designated for this safety system and is not available to the end user for assignment to alarms.




Figure 98 Optional Hydrogen System Safety System

Appendix B: NanoTrace DF-700 Series Moisture Analyzers Serial Digital Communication Protocol Programming Guide

Note: For questions about the information in this guide, contact Servomex Customer Service.

This guide assumes a user has a basic understanding of binary/hexadecimal system and a programming background. You cannot use the Hyper Terminal program in Microsoft Windows to talk to our analyzers. There is a diagnostic utility ("NTCommD.exe") available upon request to facilitate troubleshooting and to assist in programming.

Table of Contents:

- 0. Introduction**
- 1. Protocol**
- 2. Table of Commands**
- 3. RS-232/RS-485 Wiring**
- 4. Timing**
- 5. Command Description and Examples**

0. Introduction

Servomex DF digital communication protocol messages are mainly in binary format, unless stated otherwise. The main reasons for using binary format, instead of ASCII text format are:

- 1) to ensure data integrity,
- 2) to improve communication efficiency and
- 3) to support multi-drop communication in a RS-485 loop

The disadvantage is that a terminal program, like the Hyper Terminal that comes with Microsoft Windows operating system, can't communicate with a DF analyzer in a comprehensible way.

Data integrity is important particularly in a multi-drop RS-485 loop. A host computer communicates with multiple analyzers via a serially connected loop. Without a mechanism to ensure its data integrity, a host message may not reach an intended unit, may reach beyond an intended unit, may be corrupted, a reply may or may not be from an intended unit or a reply is from an intended unit but for a different command. For DF communication protocol, data integrity is ensured by using an embedded Checksum with every message, and an echoed ID and command numbers in every reply message. Checking them serves as an indicator whether a command is successful.

A NanoTrace Dual DF-760E consists of a moisture and an oxygen sensor. Each sensor must be configured with a unique 485 ID.

Some commands are backward compatible to our oxygen-only analyzers as long as the commands share the same syntax.

1. Protocol

All numbers are in binary format, unless stated otherwise. For convenience, binary is represented in hexadecimal.

11 Computer (Host) command format:

<01><ID#><Cmd><Bytes><Data>...<Data><MsChecksum><LsChecksum><CR>

where:

< > - represents one byte

01 = 01, constant preamble to identify a computer command

ID# = ID number of intended analyzer (1-255)

Cmd = Command Number

Bytes = Number of data byte (128 max)

Data...Data = Data bytes; either in binary or ASCII text format

MsChecksum = Most significant byte of message checksum (checksum includes ID# through last Data byte)

LsChecksum = Least significant byte of message checksum

CR = Carriage return - constant synchronization terminator

12 Analyzer (Slave) response format:

<02><ID#><Cmd><Bytes><Data>...<Data><MsChecksum><LsChecksum><CR>

where:

< > - represents one byte

02 = 02, constant preamble to identify an analyzer reply

ID# = ID number of analyzer (echoed from a computer command) **Cmd** = Command number (echoed from a computer command) **Bytes** = Number of data byte (128 max)

Data...Data = Data bytes

MsChecksum = Most significant byte of message checksum (checksum includes ID# through last Data byte)

LsChecksum = Least significant byte of message checksum

CR = Carriage return - constant synchronization terminator

2. Table of Commands

Notation:

fl.: IEEE floating point format.

*: additional description available. V: variable length

Command name	Cmd #	Bytes	Reply	700 only	Description
Status Request *	0	0	3		get status
Get ppb *	1	0	4 (fl)		get averaged ppb reading
Get FS	2	0	4 (fl)		get device full scale in ppm
Get Temp	3	0	4 (fl)		get averaged temp; O2: sensor, H2O: cell
Get version	7	0	4 (fl)		get firmware version
Get model *	13	0	13		get analyzer model
Get serial No	15	0	16		get unit serial number string
Get ppb in text *	24	0	10		get ppb reading in ASCII text
Aout Range *	32	V	0		get/set analog output range
Alarms *	33	V	0		get/set alarm
GSF *	34	V	0		get/set the Gas Scale Factor
Do user Cal *	40	V	0		do a user calibration
Abort cal	41	0	0		abort command 40 in progress
Accept a Cal	42	0	0		accept a manual calibration data
Watchdog Status	67	0	4 (fl)	y	get watchdog delay timer in mS = 0002 BF20 = 180000
Prepare Log *	100	V	0	y	compile a history log
Download Log *	101	V	V	y	download a history log
Get H2O Data *	102	0	15	y	get H2O data: Status/ppb/Pcell/RmTemp
System Controls *	103	4	0	y	send a system control
Write Info *	126	V	0	y	select a background gas; send a reboot request
Read Info *	127	0	V	y	get the current background gas

3. RS-232/RS485 Wiring

Parameters are 8N1, (8 bits, no parity, one stop bit).

For using the RS-232C option, connect the unit directly to an RS-232C port at the PC. The RS-485 option requires that the PC have either an internal RS-422/485 interface board or an external RS-232C to RS-485 converter.

Only three wires are required for the RS-232C interface. The RTS and CTS lines are not used.

RS-232

Analyzer COMM Connector J15 on analyzer	to Personal Computer	
pin	Female DB9 connector	Female DB25 connector
1 RXD	3	2
2 TXD	2	3
3 RTS (not used)		
4 CTS (not used)		
5 RXD- (not used)		
6 (not used)		
7 TXD- (not used)		
8 GND	5	7

It is recommended that the wire length not exceed 30 feet. Wire size can be anything larger than #26 AWG.

RS-485

For RS-485, the analyzer has been configured for a four-wire, full duplex interface. A recommended wire for RS-485 communications is #20 AWG twisted pair. Do not exceed 4000 feet. Connect the RS- 485 wiring as follows:

Analyzer COMM Connector J15 on analyzer	to RS-485 interface
1 RXD	TXD+
2 TXD	RXD+
3 RTS (not used)	
4 CTS (not used)	
5 RXD-	TXD-
6 (not used)	
7 TXD-	RXD-
8 GND	

4. Serial Digital Communication Timing

Readings (i.e. ppb, temperature, ...) in either a dual or a moisture only model, are updated about every two seconds. A PC can generally expect a reply to a command in less than one second. A two- second timeout is recommended for receiving a reply. The delay interval between consecutive commands is recommended at minimum 300 ms. If a command syntax is correct, but an analyzer is unable to perform the command either because it is invalid or an analyzer is busy, it will set the MS bit of the CMD byte (command number) in its return string. If the command syntax is incorrect (wrong Bytes, ID, Checksum etc.) an analyzer will not reply.

If a PC sends out a command at 1 Hz rate, the actual perceived communication rate to any unit in a 485 multi-unit loop is 1 Hz, independent of the number of units. This is the nature of a multi-drop 485 serial communication. A unit in a multi-drop loop has to check all the traffic (commands/replies) to know the destination of messages. That is the significance of the "**constant preamble**" and "**ID**". The preamble tells if a message originates from a PC host or an analyzer slave. The ID tells the destination.

5. Command Description and Examples

Command 0 : [Status Request](#)
 Command 1 : [Get ppb](#) Command 13 : [Get Model](#)
 Command 32 : [Analog Output Range](#)
 Command 33 : [Alarms](#)
 Command 34 : [GSF](#)
 Command 40 : [User Calibration](#)

Command 67 : [Watchdog Timer Delay](#)

Command 100: [Prepare Log](#) Command 101: [Download Log](#)
 Command 102: [Get H2O Data](#) Command 103: [Send a System Control](#)

Remote Sample selection Command 126: [Write Info](#) Command 127: [Read Info](#)

Command 0: Status Request

Command	Cmd#	Bytes	Reply	Description
Status Request	0	0	3	get status

Byte 2 bit assignment:	Byte 1 bit assignment:	Byte 0 bit assignment:
0 Unit isolated	0 NU	0 Over Range
1 Unreasonable Room Temp	1 NU	1 NU
2 Warm up	2 NU	2 Invalid Data
3 H2O error/O2 sensor off	3 Command Error	3 NU
4 Pres. not in range/No O2 sensor	4 Temperature alarm	4 alarm 1 on
5 In Calibration	5 O2 electrolyte/H2O pres. alarm	5 alarm 2 on
6 NU	6 O2 Flow Alarm/H2O sys. alarm	6 alarm 3 on
7 NU	7 O2 sensor off/H2O NU alarm	7 alarm 4 on

A set (=1) bit is True.

In Calibration: an Auto/Manual Span/Zero is in progress.

When the **Invalid Data** bit is set (=1), the ppb reading is considered not reliable. The **Invalid Data** bit is set, when any of bit 0..4 in Byte 2 is set.

The alarms 1..4 are concentration alarms.

Command Error: the last command is invalid. To clear the bit, resend a valid command.

Ex. code in hexadecimal:

Get Status from the ID5 analyzer: 0105 0000 0005 0D

where

01: host (PC) identification, indicating the message from a PC

05: analyzer ID number

00: command number, 0: Get status

00: number of data bytes in the message = 0

0005: Checksum = the byte sum of 05 0000 = 05+00+00

0D: carriage return, marking the end of message

Reply: 0205 0003 0002 0100 0B0D

where

02: slave (analyzer) identification, indicating the message from an analyzer

05: analyzer ID number

00: echoed command number from the host message

03: number of data bytes in message = 3

0002 01: 3 data bytes, indicating that bit 1 in byte 2 and bit 0 in byte 3 are set. The sensor is in isolation and in calibration.

00 0B: CheckSum = the byte sum of 05 0003 0002 01 = 05+00+03+00+02+01

0D : carriage return, marking the end of message

The ID and Command number are identical to the original host message, and Checksum passes error checking. The success and integrity of the host command and reply messages are confirmed.

Command 1: Get ppb

Command Cmd# Bytes Reply Description

Get ppb 1 0 4 (fl) get ppb reading

Ex. code in hexadecimal

Get ppb from the ID2 analyzer: 0102 0100 0003 0D

where

01: host (PC) identification, indicating the message from a PC

69: analyzer ID number

01: command number, 1: Get ppb

00: number of data bytes in the message = 0

0003: Checksum = the byte sum of 02 0100 = 02+01+00

0D: carriage return, marking the end of message

Reply: 0202 0104 3CC1 1C2B 014B 0D

where

02: slave (analyzer) identification, indicating the message from an analyzer

02: analyzer ID number

01: echoed command number from the host message

04: number of data byte in message = 4

3CC1 1C2B: 4 data bytes, = 0.024 ppb

014B: Checksum = the byte sum of **02 0104 3CC1 1C2B** = 02+01+04+3C+C1+1C+2B

0D : carriage return, marking the end of message

The ID and Command number are identical to the original host message, and the Checksum

passes error checking. The success and integrity of the host command and reply messages are confirmed.

The **3CC1 1C2B** is the representation of ppb reading in single precision IEEE floating point format.

Most programming languages have a built-in function or a type-cast for the conversion. To do it yourself:

IEEE Standard: For 32-bit system (single precision) 1-bit sign (most significant)
8-bit exponent
23-bit mantissa (least significant)
The value of an IEEE floating point is given as

$$\text{value} = (-1)^{\text{sign}} * 2^{(\text{exponent}-127)} * (1 + \text{fraction})$$

where **fraction = mantissa / 2^{23}**

3CC1 1C2B gives sign=0, exponent = **121**, mantissa = **4267051**

$$(-1)^0 * 2^{(121-127)} * (1 + 4267051/2^{23}) = 1 * 0.015625 * 1.508672 = 0.024$$

Command 13: Get Model Number

<u>Command</u>	<u>Cmd#</u>	<u>Bytes</u>	<u>Reply</u>	<u>Description</u>
Get Model	13	0	10	

Byte assignment

- 0 NU
- 1 NU
- 2 Space (Hex 20)
- 3 Space (Hex 20)
- 4 Space (Hex 20)
- 5-8 model string, UTOA (oxygen) and UTHA (moisture)
- 9 Null (Hex 0)

get Model

Command 24: Get ppb in Text

<u>Command</u>	<u>Cmd#</u>	<u>Bytes</u>	<u>Reply</u>	<u>Description</u>
Get ppb in Text	24	0	10	get ppb reading in ASCII text

Returned data bytes: 10 bytes fix
length Format: scientific notation
"**±1.XXXE±YY**" in ppb
ex. +1.234E+01

Command 32: Analog Output Range

Command	Cmd#	Bytes	Reply	Description
AOut Range	32	V	V	Set/Read analog (0-10V and 4-20mA) output range

Data Bytes in ASCII text:

"**AnalyzerModel Mode zero FullScale InCalRelay CalFreeze ExpRangeARelay ExpRangeAFS
ExpRangeBRelay ExpRangeBFS
NULL**" where

AnalyzerModel: **NTOA**: oxygen, **NTHA**: moisture

Mode: **S**: set; **R**: read settings;

for the read ('R') case the remaining parameters in the command are returned

zero: in ppm, the minimum output scale

FullScale: in ppm, the maximum output scale

InCalRelay: 0-4, 0 for no assigned relay, relay for

calibration **CalFreeze**: 0:Off; 1:On, when on, no

analog output update **ExpRangeARelay**: 0-4, 0 for

no assigned relay **ExpRangeAFS**: expanded range A

full scale in ppm **ExpRangeBRelay**: 0-4, 0 for no

assigned relay **ExpRangeBFS**: expanded range B full

scale in ppm

NULL: string terminator, NUL character

All values must be valid, or the Command Error bit is set in the Status byte. Note that for the analog outputs the **ExpRangeAFS** value must be either 0 or greater than the FS value, **ExpRangeBFS** value must be 0 or greater than **FullScale** and **ExpRangeAFS**. If **ExpRangeAFS** is zero the **ExpA** function is disabled. **ExpB** works the same way. If the ID and AnalyzerModel are inconsistent, the MS bit of the CMD byte in the reply is set.

Ex. Code:

Get **AOut Range** from an oxygen analyzer with a 485 ID 1:

Command:

Hex: 0101 2033 4E54 4841 2052 2030 2E30 3030 3030 3020 302E 3030 3030 3030 2030 2030
2030

2030 2E30 3030 3030 3020 3020 302E 3030 3030 3030

0009 A90D

ASCII: 3NTHA R 0.000000 0.000000 0 0 0 0.000000 0

0.000000

*Note: the characters after the "R" are ignored and do not have to be present.

: non-printable ASCII

Reply:

Hex: 0201 202B 302E 3030 3030 3030 2031 2E30 3030 3030 3020 3020 3020 3020 322E 3030
3030

3030 2030 2033 2E30 3030 3030 3007 EA0D

ASCII: +0.000000 1.000000 0 0 0 2.000000 0 3.000000ê

The reply
indicates

zero: 0.0

ppm

FullScale: 1.00 ppm

InCalRelay: 0

CalFreeze: 0

ExpRangeARelay: 0

ExpRangeAFS: 2.00 ppm

ExpRangeBRelay: 0

ExpRangeBFS: 3.00 ppm

Set **AOut Range** from an oxygen analyzer with a 485 ID 1 to:

zero: 0.0

ppm

FullScale: 1.0 ppm

InCalRelay: 2

CalFreeze: 0

ExpRangeARelay: 3

ExpRangeAFS: 0.30 ppm

ExpRangeBRelay: 0

ExpRangeBFS: 10.0 ppm

Command:

Hex: 0101 2033 4E54 4841 2053 2030 2E30 3030 3030 3020 312E 3230 3030 3030 2030 2030
2032

2032 2E35 3030 3030 3020 3020 352E 3030 3030 3030

0009 BB0D

ASCII: 3NTHA S 0.000000 1.200000 0 0 2 2.500000 0

5.000000

Reply:

Hex: 0201 2000

0021 0D

Command 33: Alarms

<u>Command</u>	<u>Cmd#</u>	<u>Bytes</u>	<u>Reply</u>	<u>Description</u>
Set Alarms	33	V	0	Set/Read alarm thresholds

Set Data bytes format in ASCII text:

"**AnalyzerModel S AlarmID Alarm# On/Off Relay# Audible** [**LowThreshold HighThreshold Deadband**] **NULL**"

Set Reply: no data bytes

Set Note: for 'O' alarms if the high threshold is larger than low threshold the high value is used

for setpoint and the “trip” is set to “above”. If low is larger than high the low value is used for setpoint and “trip” is set to “below”. If the ID and AnalyzerModel are inconsistent, the MS bit of the CMD byte in the reply is set.

Read Data bytes in ASCII

text: "NTOA R AlarmID
Alarm# NULL"

Read Reply Data bytes:

"On/Off Relay# Audible [LowThreshold HighThreshold Deadband] NULL "

Read Note: the remaining parameters (after ‘O’ , ‘T’ , ‘E’ or ‘F’ and the alarm #) are returned.

Where

AnalyzerModel: NTOA: oxygen, NTHA: moisture

S: set; **R:** read settings

AlarmID: 1 byte; either O, T, E or F;

O/H: concentration, 4 available alarms; O: O₂,

H:H₂O T: temperature, 1 available alarm

E: electrolyte for O₂, Pressure for H₂O, 1 available

alarm F: flow for O₂, System Error for H₂O, 1 available
alarm

Alarm#: 1 byte; 1-4 for O, and 1 for T, E and F

On/Off: 1 byte; 0 for Off, 1 for On alarm

Relay#: 1 byte; 0-4; 0 for no assigned relay; 1..4 corresponds relay 1..4 for H₂O; 5..8 for O₂

Audible: 1 byte; 0 for Off, 1 for

On for AlarmID= O or T

LowThreshold: in ppm for O; in °C

for T **HighThreshold:** in ppm for O; in

°C for T **Deadband:** in ppm for O; in

°C for T **NULL:** string terminator, NUL
character

For AlarmID = O, If **HighThreshold** > **LowThreshold**, the **HighThreshold** is the setpoint and the "Trip" is set to "Above". If **LowThreshold** > **HighThreshold**, the **LowThreshold** is the setpoint and the "Trip" is set to "Below".

Refer to [Command 32](#) for example codes

Command 34: GSF

Command Cmd# Bytes Reply Description

Set GSF **34** **V** **0** **Set/Read gas scale factor**

To oxygen Analyzer: Write format Data bytes in ASCII text: "NTOA S SampleGSF SpanGSF" Write reply: no data bytes Read format in ASCII text: "NTOA R" Read reply: "SampleGSF SpanGSF " If the ID and AnalyzerModel are inconsistent, the MS bit of the CMD byte in the reply is set.	To moisture Analyzer: Write format Data bytes in ASCII text: "NTHA S SampleGSF SpanGSF" Write reply: no data bytes Read format: "NTHA R" Read reply: "SampleGSF SpanGSF" If the ID and AnalyzerModel are inconsistent, the MS bit of the CMD byte in the reply is set.
--	---

Command 40: User Calibration

Command Cmd# Bytes Reply Description

Data Bytes:

<**Type**><**Mode**><ppm Span value in ASCII> **Type**:

0: Span, 1: Zero

Mode: 0: Manual, 1: Auto

Ex1. Auto Span at 7.05 ppm

Data Bytes in hex: **0001 372E 3035**

Ex2. Manual zero

User Cal 40 V 0 activate user calibration

Command 67: Read Watchdog Timer Delay

Data Bytes in hex: **0100 30**

<u>Command</u>	<u>Cmd#</u>	<u>Bytes</u>	<u>Reply</u>	<u>Description</u>
Send a Control	67	0	4 (fl)	Read watchdog timer delay

Data byte is a unsigned (4-byte) long integer binary format, the LSB first. The delay is hard-coded to 3 minutes.

thus reply data byte = 0002 BF20 = 180000 ms.

Command 100: Prepare Log

Command Cmd# Bytes Reply DF-760 Description

Prepare Log 100 V V P 1. Set: prepare history log for download
2. Read: status of log file preparation

Set Data bytes format in ASCII text:

"AnalyzerModel S MM/DD/YY HH:MM:SS MM/DD/YY HH:MM:SS Sec NULL"

If the ID and AnalyzerModel are inconsistent, the MS bit of the CMD byte in the

reply is set. To moisture analyzer at ID 1:

Start Date: 06/24/01

Start Hour:

15:0:0 End

Date:

06/25/01 End

hour: 15:0:0

Data interval:

60 sec

Hex: 0101 642B 4E54 4841 2053 2030 362F 3234 2F30 3120 3135 3A30 3A30 2030 362F 3235 2F30 3120

3135 3A30 3A30 2036 3020 0008 DF0D

ASCII: d+NTHA S 06/24/01 15:0:0 06/25/01 15:0:0 60

Set reply:

no data bytes

Read

format:

"AnalyzerModel R"

Hex: 0101 642B 4E54 4841 2052 2030 362F 3234 2F30 3120 3135 3A30 3A30 2030 362F 3235 2F30 3120

3135 3A30 3A30 2036 3020 0008 DE0D

ASCII: d+NTHA R 06/24/01 15:0:0 06/25/01 15:0:0 60

*Note: the characters after the "R" are ignored and do not have to be present.

Read reply:

"MM/DD/YY HH:MM:SS MM/DD/YY HH:MM:SS Sec Msg NULL"

Hex: 0201 6437 3036 2F32 342F 3031 2031 353A 3030 3A30 3020 3036 2F32 352F 3031 2031 353A 3030

3A30 3020 3630 2036 3031 3620 4259 5445 5320 5245 4144 590B F60D

ASCII: d706/24/01 15:00:00 06/25/01 15:00:00 60 6016 BYTES READY ö

Where

AnalyzerModel: NTOA for oxygen, NTHA for moisture

MM/DD/YY: log start/end date

HH:MM:SS: log start/end time

Sec: data point interval in seconds, valid value: multiple of 60 seconds

Msg: either "BUSY", "NO FILE" or "XXXX BYTES READY"

BUSY: preparing log

NO FILE: no log file for download

XXXX BYTES READY: log is ready for download

NULL: string terminator, NUL character = ASCII 0

Command 101: Download Log

<u>Command</u>	<u>Cmd#</u>	<u>Bytes</u>	<u>Reply</u>	<u>DF-760</u>	<u>Description</u>
Download Log	101	V	V	P	download 128 bytes from the prepared log file
		starting at	Offset		

Data bytes format in binary:

[4ByteOffset]

Reply Data bytes:

[size number of data bytes]

The command to download the last segment of log file:

0101 6504 0017 0000 0081 0D

The offset is = 0017 0000 in hex = 5888 in decimal. To convert the hex to decimal using a calculator, the hex offset bytes need to be reversed to 0000 1700.

Get history log data (concentration in ppm) from the start of the log file plus the 4 byte offset (msb first). Offset must be less than files size Reported by command 100. The log file starts with a 128 byte header, followed by continuous data points without a delimiter between points. A data point is presented in 4 byte [IEEE floating point format](#). The header details data information, such as origin, timing, etc.

If the ID and AnalyzerModel are inconsistent, the MS bit of the CMD byte in the

reply is set. The sample header and first two data points in ASCII :

s/n: ND-00002, type: H2O, start: 6/24/01 3:00 PM, end: 6/25/01 3:00 PM, min/pt: 1, Npts: 1441, Nblks

(incl hdr): 47

06/24/2001 15:00:00 1.375E+1

06/24/2001 15:01:00 1.399E+1

The number after date/time stamp is in ppb.

Command 102: Get H2O Data

<u>Command</u>	<u>Cmd#</u>	<u>Bytes</u>	<u>Reply</u>	<u>DF-760</u>	<u>Description</u>
Get H2O Data	102	0	15	P	H2O Data: Status_Bytes, ppb, Pcell and Room_Temperature

Reply Data bytes:

[Status_Bytes (3 bytes), ppb (4 bytes), Pcell (4 bytes) and Room_Temperature (4 bytes)]

Data:

Status_Bytes: same as the Command 1 "Status Request"

ppb: same as the Command 1 "get ppb"

Room_Temperature: the temperature in Celsius inside the chassis

Pcell: cell pressure in torr

Command 103: Send a System Control

<u>Command</u>	<u>Cmd#</u>	<u>Bytes</u>	<u>Reply</u>	<u>Description</u>
Send a Control	103	4	0	Send a control

Data byte is a unsigned (4-byte) long integer binary format, the LSB

first. Data byte number assignment in decimal (hexadecimal):

- 65 (0x41): Isolate a sensors (ID dependant)
- 66 (0x42): Restore a sensors Process gas Flow (ID dependant)
- 67 (0x43): Isolate System (ID independent)
- 68 (0x44): Restore System Process Gas Flow (ID independent)
- 70 (0x46): Turn off Oxygen sensor (ID independent)
- 71 (0x47): Turn On oxygen sensor (ID independent)

Ex. code in hexadecimal

Isolate System: 0102 6704 4300 0000 00B0 0D

where

01: host (PC) identification, indicating the message from a PC

02: analyzer ID number

67: command number, 103: Send a

control **04:** number of data bytes in

the message = 0 **4300 0000:** 4 data

bytes LSB first

00B0: Checksum = the byte sum of 02 6704 4300 0000 = 02+67+04+43

0D: carriage return, marking the end of message

The example code isolates both H2O and O2 sensors, if so equipped.

Command 126: Write Info

<u>Command</u>	<u>Cmd#</u>	<u>Bytes</u>	<u>Reply</u>	<u>Description</u>
Send a Control	126	V	0	Write Info

Functionality: depending on the data byte parameter.

Timing

A new sample name selected via Command 126 will be displayed on the front screen in about 3 seconds. Similarly in about 3 seconds, the new name is available for Command 127 reply.

----- 2017-10-23 background sample gas selection-----

Command 126 (HS): Select background sample gas

write info, data bytes are case sensitive for all models, except

745sgMax. To select a background sample gas or a template:

"HS gas_name" or "HS #template".

where HS: function identifier, following H2O background sample name, #: prefix for a template name

Usage: ex. "HS Ar" or "HS

#AIR" Return data byte: 0

Only 745sgMax supports sample mixing via a template . A user-generated template must already exist in the analyzer.

For non-745sgMax, the command is case sensitive and a gas name must be exact such as He, O2, H2, He, CO2, Ar, HCl, or NH3.

"Remote Sample Selection": write selection (see Cmd 127 to read)

----- as of firmware v6.3.7, 2019-04-22 -----

Command 126 (RB 101): Select background sample gas

write Remote reboot command with : "**RB 101**"

reply Data bytes: if command ok, various with arbitrary data

Command 127: Read Info

<u>Command</u>	<u>Cmd#</u>	<u>Bytes</u>	<u>Reply</u>	<u>Description</u>
Send a Control	127	0	v	Read Info

Functionality: depending on the data byte parameter.

"Remote Sample Selection": read selection (see Cmd 126 to write)

Timing

A new sample name selected via Command 126 will be displayed on the front screen in about 3 seconds. Similarly in about 3 seconds, the new name is available for Command 127 reply.

Command 127: read info

To read back the current sample or template name:

"HS": data bytes optional

Usage: "HS" or ""

Return data bytes: various either "gas or template_name" or if previously erroneous Cmd 126, "NOK, gas/template name".